



**JAWAHARLAL INSTITUTE OF POST GRADUATE MEDICAL
EDUCATION & RESEARCH (JIPMER)**



GOVERNMENT OF INDIA

(An Institution of National Importance under Ministry of Health & Family Welfare)
Dhanvantri Nagar, Puducherry-605006

**OPEN e-TENDER
ENQUIRY DOCUMENT
FOR
RATE CONTRACT
FOR
SUPPLY OF CONSUMABLES/NON-CONSUMABLES ITEMS
FOR
DEPARTMENT OF CSSD
for a period of ONE Year**

(This document consisting of **46** pages)

**JIPMER,
DHANVANTARI NAGAR,
PUDUCHERRY 605006**

**Phone: - 0413-2296016/15/47 ; Fax:- 0413-2272067/66
0413-2296602**

Email: jipmercssdstore@gmail.com URL: - www.jipmer.edu.in

INDEX

Section	Topic	Page No.
Section-I	Notice inviting Tender (NIT)	03
Section-II	Schedule of Requirements and Specifications	05
Section-III	General Instructions to Tenderers (GIT)	25
Section-IV	General Conditions of Contract (GCC)	38
Section-V	Bid Application Form	40
Section-VI	Manufacturer's Authorization Form	41
Section-VII	Tender Form	42
Section-VIII	Check list for Submission of Tender	43
Appendix-A	Self-certification format for claiming purchase preference under the "Public Procurement preference to Make in India" order	44
Section-IX	GFR-144 (xi) compliance certificate	46

SECTION-I

JAWAHARLAL INSTITUTE OF POSTGRADUATE MEDICAL EDUCATION AND RESEARCH PUDUCHERRY – 6

(Institution of National Importance under the Ministry of Health & Family Welfare,
Govt. of India)

PURCHASE SECTION

No.JIP/Pur.5(3)/OeT/CSSD/2024-25

Dated: 08.07.2024

NOTICE INVITING e-TENDER (e-NIT)

Sub: Supply of Consumables/Non-Consumables Items

e-Tenders are invited from eligible and qualified domestic manufacturers or their distributors for the RATE CONTRACT for supply of CONSUMABLES/NON-CONSUMABLES ITEMS for the Department of CSSD for a period of ONE Year.

1. Scope of work: Supply of consumables for the Department of CSSD
2. Tender timelines:
 - i. Opening date & time for download of Tender document: **06.00 PM** on **08.07.2024**
 - ii. Last date for receipt of pre-bid queries: **06.00 PM** on **11.07.2024**
 - iii. Pre-bid queries regarding Items and Samples can be made through e-mail to:
jipmercssidstore@gmail.com and other Document related queries may be raised through purchase.jipmer@gmail.com
 - iv. Opening date & time for submission of online bids: **09.00 AM** on **12.07.2024**
 - v. Closing date & time for submission of online bids **06.00 PM** on **05.08.2024**
 - vi. Date and Time of opening of online bids for Technical evaluation: **06.00 PM** on **06.08.2024**
 - vii. Date & time of opening of Price Bid: To be intimated later.
3. Earnest Money Deposit: **Rs.5,000/-** to be paid through SBI collect only. The EMD shall be returned without interest to the non-successful tenderers after acceptance of award of contract by the successful bidder.
4. Tender Processing Fee (Non-refundable): **Rs. 590/- (including 18% GST) shall be paid through SBI collect only.**
5. Interested bidders are advised to download the complete Tender Enquiry document from the websites www.jipmer.edu.in or <https://eprocure.gov.in/eprocure/app> for complete details.

6. The prospective bidders must register with the E-procurement system of <https://eprocure.gov.in/eprocure/app>. Special Instructions to the bidders for the e-submission of the bids online through this e-Procurement Portal on completion of the registration process is given in <https://eprocure.gov.in/eprocure/app>, the bidders will be provided user ID and password upon enrollment. In order to submit the bids electronically, bidders are required to have a valid Class 3 Digital Signature Certificate (signing and encryption/ decryption certificates).
7. Bidders are requested to read the bidders help document on e-tender web site link before proceeding for bidding.
8. Post receipt of User ID & Password, Bidders can log on for downloading & uploading tender document.
9. The bidders shall submit the required EMD (as per G.I.T clause 2) before the due date and time mentioned above.
10. The online submission of tender(s) can only be done through <https://eprocure.gov.in/eprocure/app>
11. Bidders shall ensure that their tender(s), complete in all respects, are submitted online through <https://eprocure.gov.in/eprocure/app> e-portal (as described above) only.
12. Prospective bidders are advised to browse the above websites regularly before submission of their bids as any further amendments will be published in these websites only.

Asst. Officer in Charge
Purchase Section,
For Director,
JIPMER, Puducherry

Section-II

SCHEDULE OF REQUIREMENTS AND SPECIFICATIONS

LIST OF ITEMS REQUIRED

1. Scope of work : Supply of Consumables/Non-Consumables Items for the department of **CSSD** as detailed below

Sl. No.	Description of Item	Quantity	Units
1	CONSUMABLES ITEMS		
1.001	<p>Surgical Gloves for General routine operations / ICU usage: Size - 6.0: Single use, sterile, disposable, polimer coated powder free, natural latex surgical gloves with finish of smooth inner surface and rough outer surface used for general routine operations. Gloves Should have high level of tactile sensitivity, secure grip, good resistant to tear, ease of donning and reduction of reflective glare. Packing: Number of gloves in a pack: 2 Nos. (1 pair), outer packing should be poly pouch or lacquer coated peel down paper or blister poly peel. Inner packing with medical grade paper. Packed to maintain sterility during shipping & storage and permit opening without contamination of the gloves. Colour: Translucent. Dimensions: Size(Number): 6 Thickness: 0.20 - 0.24 mm. Self life: 5 years. The product should have at least 2/3 rd of the total shelf life available at the time of supply. Standards: BIS certified, confirming IS 4148: 1989 (2021) / IS 13422: 1992 (2018) or equivalent standards CE / USFDA approved. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL accredited Lab to prove conformity to the specifications.</p>	125000	pair
1.002	<p>Surgical Gloves for General routine operations / ICU usage: Size 6.5 Single use, sterile, disposable, polimer coated powder free, natural latex surgical gloves with finish of smooth inner surface and rough outer surface used for general routine operations. Gloves Should have high level of tactile sensitivity, secure grip, good resistant to tear, ease of donning and reduction of reflective glare. Packing: Number of gloves in a pack: 2 Nos. (1 pair), outer packing should be poly pouch or lacquer coated peel down paper or blister poly peel. Inner packing with medical grade paper. Packed to maintain sterility during shipping & storage and permit opening without contamination of the gloves. Colour: Translucent. Dimensions: Size(Number): 6.5 Thickness: 0.20 - 0.24 mm. Self life: 5 years. The product should have at least 2/3 rd of the total shelf life available at the time of supply. Standards: BIS certified, confirming IS 4148: 1989 (2021) / IS 13422: 1992 (2018) or equivalent standards CE / USFDA approved. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL accredited Lab to prove conformity to the specifications.</p>	250000	pair
1.003	<p>Surgical Gloves for General routine operations / ICU usage: Size 7.0 : Single use, sterile, disposable, polimer coated powder free, natural latex surgical gloves with finish of smooth inner surface and rough outer surface used for general routine operations. Gloves Should have high level of tactile sensitivity, secure grip, good resistant to tear, ease of donning and reduction of reflective glare. Packing: Number of gloves in a pack: 2 Nos. (1 pair), outer packing should be poly pouch or lacquer coated peel down paper or blister poly peel. Inner packing with medical grade paper. Packed to maintain sterility during shipping & storage and permit opening without contamination of the gloves. Colour: Translucent. Dimensions: Size(Number): 7 Thickness: 0.20 - 0.24 mm. Self life: 5 years. The product should have at least 2/3 rd of the total shelf life available at the time of supply. Standards: BIS certified, confirming IS 4148: 1989 (2021) / IS 13422: 1992 (2018) or equivalent standards CE / USFDA approved. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL accredited Lab to prove conformity to the specifications.</p>	450000	pair

1.004	<p>Surgical Gloves for General routine operations / ICU usage: Size 7.5 Single use, sterile, disposable, polimer coated powder free, natural latex surgical gloves with finish of smooth inner surface and rough outer surface used for general routine operations. Gloves Should have high level of tactile sensitivity, secure grip, good resistant to tear, ease of donning and reduction of reflective glare. Packing: Number of gloves in a pack: 2 Nos. (1 pair), outer packing should be poly pouch or lacquer coated peel down paper or blister poly peel. Inner packing with medical grade paper. Packed to maintain sterility during shipping & storage and permit opening without contamination of the gloves. Colour: Translucent. Dimensions: Size(Number): 7.5 Thickness: 0.20 - 0.24 mm. Self life: 5 years. The product should have at least 2/3 rd of the total shelf life available at the time of supply. Standards: BIS certified, confirming IS 4148: 1989 (2021) / IS 13422: 1992 (2018) or equivalent standards CE / USFDA approved. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL acredited Lab to prove conformity to the specifications.</p>	450000	pair
1.005	<p>Surgical Gloves for General routine operations / ICU usage: Size 8.0 : Single use, sterile, disposable, polimer coated powder free, natural latex surgical gloves with finish of smooth inner surface and rough outer surface used for general routine operations. Gloves Should have high level of tactile sensitivity, secure grip, good resistant to tear, ease of donning and reduction of reflective glare. Packing: Number of gloves in a pack: 2 Nos. (1 pair), outer packing should be poly pouch or lacquer coated peel down paper or blister poly peel. Inner packing with medical grade paper. Packed to maintain sterility during shipping & storage and permit opening without contamination of the gloves. Colour: Translucent. Dimensions: Size(Number): 8 Thickness: 0.20 - 0.24 mm. Self life: 5 years. The product should have at least 2/3 rd of the total shelf life available at the time of supply. Standards: BIS certified, confirming IS 4148: 1989 (2021) / IS 13422: 1992 (2018) or equivalent standards CE / USFDA approved. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL acredited Lab to prove conformity to the specifications.</p>	200000	pair
1.006	<p>Surgical Gloves for Speciality surgeries: Size 6.0: Single use, sterile, disposable, polimer coated powder free, natural latex surgical gloves with finish of smooth inner surface and rough outer surface, tight beaded cuff, used for speciality surgeries. Textured finish: Gloves should ensure secure grip for confident instrument handling under dry or damp conditions and reduction of reflective glare. Optimized thickness: Gloves should have ideal balance of tactile sensitivity and protection, shape of curved fingers thumb ball, rounded finger tips, to reduce hand fatigue Donning technology: Gloves should have internal coating to facilitate effortless / easy donning and doffing with either dry or damp hands. Chemo tested: Gloves Should be tested and safe to use with wide range of drugs. Packing: Number of gloves in a pack: 2 Nos.(1 pair), outer packing should be poly pouch or lacquer coated peel down paper or blister poly peel. Inner packing with medical grade paper. Packed to maintain sterility during shipping & storage and permit opening without contamination of the gloves. Colour: Translucent / Brown / Green to reduce reflective glare. Dimensions: Size(Number): 6. Self life: 5 years. The product should have at least 2/3 rd of the total shelf life available at the time of supply. Standards: BIS certified or ASTM D3577/D7160 or EN 16523-1/EN 374-2 AND -4/EN 4551-4 or EN ISO 21420:2020/ISO 10282. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility or certificates to be available in the OEM website .</p>	125000	pair

1.007	<p>Surgical Gloves for Speciality surgeries: Size 6.5: Single use, sterile, disposable, polimer coated powder free, natural latex surgical gloves with finish of smooth inner surface and rough outer surface, tight beaded cuff, used for speciality surgeries. Textured finish: Gloves should ensure secure grip for confident instrument handling under dry or damp conditions and reduction of reflective glare. Optimized thickness: Gloves should have ideal balance of tactile sensitivity and protection, shape of curved fingers thumb ball, rounded finger tips, to reduce hand fatigue Donning technology: Gloves should have internal coating to facilitate effortless / easy donning and doffing with either dry or damp hands. Chemo tested: Gloves Should be tested and safe to use with wide range of drugs. Packing: Number of gloves in a pack: 2 Nos.(1 pair), outer packing should be poly pouch or lacquer coated peel down paper or blister poly peel. Inner packing with medical grade paper. Packed to maintain sterility during shipping & storage and permit opening without contamination of the gloves. Colour: Translucent / Brown / Green to reduce reflective glare. Dimensions: Size(Number): 6.5. Self life: 5 years. The product should have at least 2/3 rd of the total shelf life available at the time of supply. Standards: BIS certified or ASTM D3577/D7160 or EN 16523-1/EN 374-2 AND -4/EN 4551-4 or EN ISO 21420:2020/ISO 10282. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility or certificates to be available in the OEM website .</p>	250000	pair
1.008	<p>Surgical Gloves for Speciality surgeries: size 7.0 : Single use, sterile, disposable, polimer coated powder free, natural latex surgical gloves with finish of smooth inner surface and rough outer surface, tight beaded cuff, used for speciality surgeries. Textured finish: Gloves should ensure secure grip for confident instrument handling under dry or damp conditions and reduction of reflective glare. Optimized thickness: Gloves should have ideal balance of tactile sensitivity and protection, shape of curved fingers thumb ball, rounded finger tips, to reduce hand fatigue Donning technology: Gloves should have internal coating to facilitate effortless / easy donning and doffing with either dry or damp hands. Chemo tested: Gloves Should be tested and safe to use with wide range of drugs. Packing: Number of gloves in a pack: 2 Nos.(1 pair), outer packing should be poly pouch or lacquer coated peel down paper or blister poly peel. Inner packing with medical grade paper. Packed to maintain sterility during shipping & storage and permit opening without contamination of the gloves. Colour: Translucent / Brown / Green to reduce reflective glare. Dimensions: Size(Number): 7. Self life: 5 years. The product should have at least 2/3 rd of the total shelf life available at the time of supply. Standards: BIS certified or ASTM D3577/D7160 or EN 16523-1/EN 374-2 AND -4/EN 4551-4 or EN ISO 21420:2020/ISO 10282. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility or certificates to be available in the OEM website .</p>	450000	pair
1.009	<p>Surgical Gloves for Speciality surgeries: size 7.5: Single use, sterile, disposable, polimer coated powder free, natural latex surgical gloves with finish of smooth inner surface and rough outer surface, tight beaded cuff, used for speciality surgeries. Textured finish: Gloves should ensure secure grip for confident instrument handling under dry or damp conditions and reduction of reflective glare. Optimized thickness: Gloves should have ideal balance of tactile sensitivity and protection, shape of curved fingers thumb ball, rounded finger tips, to reduce hand fatigue Donning technology: Gloves should have internal coating to facilitate effortless / easy donning and doffing with either dry or damp hands. Chemo tested: Gloves Should be tested and safe to use with wide range of drugs. Packing: Number of gloves in a pack: 2 Nos.(1 pair), outer packing should be poly pouch or lacquer coated peel down paper or blister poly peel. Inner packing with medical grade paper. Packed to maintain sterility during shipping & storage and permit opening without contamination of the gloves. Colour: Translucent / Brown / Green to reduce reflective glare. Dimensions: Size(Number): 7.5. Self life: 5 years. The product should have at least 2/3 rd of the total shelf life available at the time of supply. Standards: BIS certified or ASTM D3577/D7160 or EN 16523-1/EN 374-2 AND -4/EN 4551-4 or EN ISO 21420:2020/ISO 10282. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility or certificates to be available in the OEM website .</p>	450000	pair

1.010	<p>Surgical Gloves for Speciality surgeries: size 8.0: Single use, sterile, disposable, polimer coated powder free, natural latex surgical gloves with finish of smooth inner surface and rough outer surface, tight beaded cuff, used for speciality surgeries. Textured finish: Gloves should ensure secure grip for confident instrument handling under dry or damp conditions and reduction of reflective glare. Optimized thickness: Gloves should have ideal balance of tactile sensitivity and protection, shape of curved fingers thumb ball, rounded finger tips, to reduce hand fatigue Donning technology: Gloves should have internal coating to facilitate effortless / easy donning and doffing with either dry or damp hands. Chemo tested: Gloves Should be tested and safe to use with wide range of drugs. Packing: Number of gloves in a pack: 2 Nos.(1 pair), outer packing should be poly pouch or lacquer coated peel down paper or blister poly peel. Inner packing with medical grade paper. Packed to maintain sterility during shipping & storage and permit opening without contamination of the gloves. Colour: Translucent / Brown / Green to reduce reflective glare. Dimensions: Size(Number): 8. Self life: 5 years. The product should have at least 2/3 rd of the total shelf life available at the time of supply. Standards: BIS certified or ASTM D3577/D7160 or EN 16523-1/EN 374-2 AND -4/EN 4551-4 or EN ISO 21420:2020/ISO 10282. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility or certificates to be available in the OEM website .</p>	200000	pair
1.011	<p>Sterile Examination Gloves Specification: Size: Small (5"-6") 1) Type of Glove : Single use, sterile, disposable, polimer coated powder free, natural latex gloves for medical examination . Gloves should not be sticky and good resistant to tear as per standard 2) Size: Small(5"-6"), 3)Packing for gloves: Individual packing for single sterile glove within the medical grade paper with outer cover. Number of glove in a pack: single 4) Certification: BIS certified confirming with IS:15354, ISI marked. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL acredited Lab to prove conformity to the specifications.</p>	500000	piece
1.012	<p>Sterile Examination Gloves Specification: Size: Medium (7"-8") 1) Type of Glove : Single use, sterile, disposable, polimer coated powder free, natural latex gloves for medical examination . Gloves should not be sticky and good resistant to tear as per standard 2) Size: Medium(7"-8") 3)Packing for gloves: Individual packing for single sterile glove within the medical grade paper with outer cover. Number of glove in a pack: single 4) Certification: BIS certified confirming with IS:15354, ISI marked. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL acredited Lab to prove conformity to the specifications.</p>	1000000	piece
1.013	<p>Sterile Examination Gloves Specification: Size: Large (8.5"- 9.5") 1) Type of Glove : Single use, sterile, disposable, polimer coated powder free, natural latex gloves for medical examination . Gloves should not be sticky and good resistant to tear as per standard 2) Size: Large(8.5"-9.5") 3)Packing for gloves: Individual packing for single sterile glove within the medical grade paper with outer cover. Number of glove in a pack: single 4) Certification: BIS certified confirming with IS:15354, ISI marked. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL acredited Lab to prove conformity to the specifications.</p>	500000	piece
1.014	<p>Unsterile Examination Gloves Specifications: Size: Small (5"-6") 1)Type of Glove : Single use, unsterile, disposable, polimer coated powder free, natural latex gloves for medical examination . Gloves should have good resistant to tear as per standard and not to be sticky with each other. 2)Size: Small(5"-6") 3) Packing: Multi units in one packet: 100 gloves. Certification: BIS certified confirming with IS:15354, ISI marked. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL acredited Lab to prove conformity to the specifications.</p>	500000	piece
1.015	<p>Unsterile Examination Gloves Specifications: Size: Medium (7"-8") 1)Type of Glove : Single use, unsterile, disposable, polimer coated powder free, natural latex gloves for medical examination . Gloves should have good resistant to tear as per standard and not to be sticky with each other. 2)Size: Medium(7"- 8") 3) Packing: Multi units in one packet: 100 gloves. Certification: BIS certified confirming with IS:15354, ISI marked. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL acredited Lab to prove conformity to the specifications.</p>	1000000	piece

1.016	Unsterile Examination Gloves Specifications:Size: Large (8.5"- 9.5") 1)Type of Glove : Single use, unsterile, disposable, polimer coated powder free, natural latex gloves for medical examination . Gloves should have good resistant to tear as per standard and not to be sticky with each other. 2)Size: Large(8.5"- 9.5") 3) Packing: Multi units in one packet: 100 gloves. Certification: BIS certified confirming with IS:15354, ISI marked. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), test reports / proof of sterility from CDSCO / Central Govt./NABL accredited Lab to prove conformity to the specifications.	500000	piece
1.017	Nitrile Gloves for medical examination: Size: Small (5"- 6") Type of gloves: Single use, unsterile, latex free, odor free, beaded cuff, textured fingers, consistent grip and high elasticity, powder-free nitrile gloves for medical examination. Gloves should have good resistant to tear as per standard and not to be sticky with each other. Pack size: 100 units. Colour: Blue. Medical level standard: ASTM 6319. OEM should produce data sheet, certificate of standards, valid licenses, manufacturing units certificates, test reports from CDSCO / Central Govt/NABL accredited Lab to prove conformity to the specifications.	500000	piece
1.018	Nitrile Gloves for medical examination: Size: Medium(7"- 8") Type of gloves: Single use, unsterile, latex free, odor free, beaded cuff, textured fingers, consistent grip and high elasticity, powder-free nitrile gloves for medical examination. Gloves should have good resistant to tear as per standard and not to be sticky with each other. Pack size: 100 units. Colour: Blue. Medical level standard: ASTM 6319. OEM should produce data sheet, certificate of standards, valid licenses, manufacturing units certificates, test reports from CDSCO / Central Govt/NABL accredited Lab to prove conformity to the specifications.	1000000	piece
1.019	Nitrile Gloves for medical examination: Size: Large (8.5"- 9.5") Type of gloves: Single use, unsterile, latex free, odor free, beaded cuff, textured fingers, consistent grip and high elasticity, powder-free nitrile gloves for medical examination. Gloves should have good resistant to tear as per standard and not to be sticky with each other. Pack size: 100 units. Colour: Blue. Medical level standard: ASTM 6319. OEM should produce data sheet, certificate of standards, valid licenses, manufacturing units certificates, test reports from CDSCO / Central Govt/NABL accredited Lab to prove conformity to the specifications.	500000	piece
1.020	Non-PVC Glove Specifications :- Disposable Non-PVC HDPE gloves of 45 G thickness. Transparent, Universal fitting size, soft & flexible, latex free, powder free.	200000	piece
1.021	Reusable nitrile chemical gloves for cleaning purpose: Size: Small (Palm with: 7cm) Reusable nitrile gloves made from Acrylonitrile Butadiene Rubber which is mechanical resistant (EN 388) to puncture, tear, blade cut, abrasion and penetration resistant (EN 374- part- 2 & 3) against chemicals & micro organisms. Grip: Gloves should have good gripping pattern palm in both wet and dry conditions. Color: Any. No of Gloves in a pack : 2 (one pair)	20000	pair
1.022	Reusable nitrile chemical gloves for cleaning purpose: Size: Medium (Palm with: 8cm) Reusable nitrile gloves made from Acrylonitrile Butadiene Rubber which is mechanical resistant (EN 388) to puncture, tear, blade cut, abrasion and penetration resistant (EN 374- part- 2 & 3) against chemicals & micro organisms. Grip: Gloves should have good gripping pattern palm in both wet and dry conditions. Color: Any. No of Gloves in a pack : 2 (one pair)	20000	pair
1.023	Reusable nitrile chemical gloves for cleaning purpose: Size: Large(Palm with: 9cm) Reusable nitrile gloves made from Acrylonitrile Butadiene Rubber which is mechanical resistant (EN 388) to puncture, tear, blade cut, abrasion and penetration resistant (EN 374- part- 2 & 3) against chemicals & micro organisms. Grip: Gloves should have good gripping pattern palm in both wet and dry conditions. Color: Any. No of Gloves in a pack : 2 (one pair)	10000	pair

1.024	<p>Sterile Disposable syringe for single use without needle 1 ml: Specifications: Syringes should be with clear syringe barrel, precise graduation with permanent marking, easy readability, Safe plunger backstop, smooth finishing of barrel & plunger. Syringe capacity: 1ml. Graduation: Unit Scale U-40 or both U-40 and ml printed in same syringe. Syringe Material: medical grade plastic (Polyethylene (PE)/ polypropylene (PP)/ polystyrene (PS). The materials used for fabrication of syringes (including lubricant) and packaging should be non toxic, do not adversely affect the efficacy, safety and acceptability. Syringe Barrel: sufficiently transparent to allow easy measurement of the volume contained in the syringe and detection of air bubbles. Length with a maximum usable capacity of at least 10% more than the nominal capacity.</p> <p>Standards: Indian standard IS 10258 -2002 latest, equivalent to ISO 7886 part 3.</p> <p>Barrel Nozzle: LUER SLIP type. Sterility & Product and packaging: Sterile, ethylene oxide sterilisation and compiles to BS EN 550:1994-ISO 11135-1:2007 or its equivalent Indian Standards)</p> <p>Primary packaging: Each syringe packed in an individual sterilized peel-pack / blister pack / ribbon pack. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	800000	piece
1.025	<p>Sterile Disposable syringe for single use without needle 2 ml: Specifications: Syringes should be with clear syringe barrel, precise graduation in ml, scale interval 0.1ml line increment according to ISO standard with permanent marking, easy readability, Safe plunger backstop, smooth finishing of barrel & plunger. Syringe capacity: 2 ml. Syringe Material: medical grade plastic (Polyethylene (PE)/ polypropylene (PP)/ polystyrene (PS). The materials used for fabrication of syringes (including lubricant) and packaging should be non toxic, do not adversely affect the efficacy, safety and acceptability. Syringe Barrel: sufficiently transparent to allow easy measurement of the volume contained in the syringe and detection of air bubbles. Length with a maximum usable capacity of at least 10% more than the nominal capacity.</p> <p>Standards: BIS certified, should be comply Indian standard IS 10258 -2002 latest, equivalent to ISO 7886 part 3. Barrel Nozzle: LUER SLIP type. Sterility & Product and packaging: Sterile, ethylene oxide sterilisation and compiles to BS EN 550:1994-ISO 11135-1:2007 or its equivalent Indian Standards)</p> <p>Primary packaging: Each syringe packed in an individual sterilized peel-pack / blister pack / ribbon pack. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	200000	piece
1.026	<p>Sterile Disposable syringe for single use without needle 5 ml:Specifications: Syringes should be with clear syringe barrel, precise graduation in ml scale interval 0.2ml line increment according to ISO standard with permanent marking, easy readability, Safe plunger backstop, smooth finishing of barrel & plunger. Syringe capacity: 5 ml Syringe Material: medical grade plastic (Polyethylene (PE)/ polypropylene (PP)/ polystyrene (PS). The materials used for fabrication of syringes (including lubricant) and packaging should be non toxic, do not adversely affect the efficacy, safety and acceptability. Syringe Barrel: sufficiently transparent to allow easy measurement of the volume contained in the syringe and detection of air bubbles. Length with a maximum usable capacity of at least 10% more than the nominal capacity.</p> <p>Standards: Indian standard IS 10258 -2002 latest, equivalent to ISO 7886 part 3.</p> <p>Barrel Nozzle: LUER SLIP type. Sterility & Product and packaging: Sterile, ethylene oxide sterilisation and compiles to BS EN 550:1994-ISO 11135-1:2007 or its equivalent Indian Standards)</p> <p>Primary packaging: Each syringe packed in an individual sterilized peel-pack / blister pack / ribbon pack. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	1600000	piece

1.027	<p>Sterile Disposable syringe for single use without needle 10 ml: Specifications: Syringes should be with clear syringe barrel, precise graduation in ml scale interval 0.2ml line increment according to ISO standard with permanent marking, easy readability, Safe plunger backstop, smooth finishing of barrel & plunger. Syringe capacity: 10 ml Syringe Material: medical grade plastic (Polyethylene (PE)/ polypropylene (PP)/ polystyrene (PS). The materials used for fabrication of syringes (including lubricant) and packaging should be non toxic, do not adversely affect the efficacy, safety and acceptability. Syringe Barrel: sufficiently transparent to allow easy measurement of the volume contained in the syringe and detection of air bubbles. Length with a maximum usable capacity of at least 10% more than the nominal capacity. Standards: Indian standard IS 10258 -2002 latest, equivalent to ISO 7886 part 3. Barrel Nozzle: LUER SLIP type. Sterility & Product and packaging: Sterile, ethylene oxide sterilisation and compiles to BS EN 550:1994-ISO 11135-1:2007 or its equivalent Indian Standards) Primary packaging: Each syringe packed in an individual sterilized peel-pack / blister pack / ribbon pack. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	1400000	piece
1.028	<p>Sterile Disposable syringe for single use without needle 20 ml: Specifications: Syringes should be with clear syringe barrel, precise graduation in ml scale interval 1ml line increment according to ISO standard with permanent marking, easy readability, Safe plunger backstop, smooth finishing of barrel & plunger. Syringe capacity: 20 ml. Syringe Material: medical grade plastic (Polyethylene (PE)/ polypropylene (PP)/ polystyrene (PS). The materials used for fabrication of syringes (including lubricant) and packaging should be non toxic, do not adversely affect the efficacy, safety and acceptability. Syringe Barrel: sufficiently transparent to allow easy measurement of the volume contained in the syringe and detection of air bubbles. Length with a maximum usable capacity of at least 10% more than the nominal capacity. Standards: Indian standard IS 10258 -2002 latest, equivalent to ISO 7886 part 3. Barrel Nozzle: LUER SLIP type. Sterility & Product and packaging: Sterile, ethylene oxide sterilisation and compiles to BS EN 550:1994-ISO 11135-1:2007 or its equivalent Indian Standards) Primary packaging: Each syringe packed in an individual sterilized peel-pack / blister pack / ribbon pack. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	400000	piece
1.029	<p>Sterile Disposable syringe for single use without needle 50 ml: Specifications: Syringes should be with clear syringe barrel, precise graduation in ml scale interval 1ml line increment according to ISO standard with permanent marking, easy readability, Safe plunger backstop, smooth finishing of barrel & plunger. Syringe capacity: 50 ml. Syringe Material: medical grade plastic (Polyethylene (PE)/ polypropylene (PP)/ polystyrene (PS). The materials used for fabrication of syringes (including lubricant) and packaging should be non toxic, do not adversely affect the efficacy, safety and acceptability. Syringe Barrel: sufficiently transparent to allow easy measurement of the volume contained in the syringe and detection of air bubbles. Length with a maximum usable capacity of at least 10% more than the nominal capacity. Standards: Indian standard IS 10258 -2002 latest, equivalent to ISO 7886 part 3. Barrel Nozzle: LUER SLIP type. Sterility & Product and packaging: Sterile, ethylene oxide sterilisation and compiles to BS EN 550:1994-ISO 11135-1:2007 or its equivalent Indian Standards) Primary packaging: Each syringe packed in an individual sterilized peel-pack / blister pack / ribbon pack. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	250000	piece

1.030	<p>Insulin Syringe: Syringe Unit scale (Units): U-40 Insulin Syringe type: As per ISO 8537:2016 or its equivalent Indian standards. Syringe with needle and packaged in a unit packaging. Syringe Barrel: sufficiently clear to enable dosages to be read and for air bubbles to be seen without difficulty. Barrel length with a usable capacity of either 10 % more than the nominal capacity. Syringe Material: medical grade plastic (Polyethylene (PE)/ polypropylene (PP)/ polystyrene (PS)). The materials used for fabrication of syringes (including lubricant) and packaging do not adversely affect the efficacy, safety and acceptability of insulin preparations</p>	300000	piece
1.031	<p>Actively Disabling Syringes without needle Specifications : - Capacity =5 ml Sterile, disposable and without needle. Actively disabling - RUP Syringe with Reuse prevention feature. 2 part syringe. Actively disabling Reuse prevention feature, permanently disable by user after completion of intended use even after multiple plunger movements to prevent subsequent reuse of the syringe, equivalent to ISO 7864, type 2-B.</p>	1400000	piece
1.032	<p>Actively Disabling Syringes without needle Specifications : - Capacity =10 ml Sterile, disposable and without needle. Actively disabling - RUP Syringe with Reuse prevention feature. 2 part syringe. Actively disabling Reuse prevention feature, permanently disable by user after completion of intended use even after multiple plunger movements to prevent subsequent reuse of the syringe, equivalent to ISO 7864, type 2-B.</p>	1400000	piece
1.033	<p>Sterile Disposable syringe for single use without needle 20 ml Luer Lock Type: Specifications: Syringes should be with clear syringe barrel, precise graduation in ml with permanent marking, easy readability, Safe plunger backstop. Syringe capacity: 20 ml Syringe Material: medical grade plastic (Polyethylene (PE)/ polypropylene (PP)/ polystyrene (PS)). The materials used for fabrication of syringes (including lubricant) and packaging should be non toxic, do not adversely affect the efficacy, safety and acceptability. Syringe Barrel: sufficiently transparent to allow easy measurement of the volume contained in the syringe and detection of air bubbles. Length with a maximum usable capacity of at least 10% more than the nominal capacity. Standards: Indian standard IS 10258 -2002 latest, equivalent to ISO 7886 part 3. Barrel Nozzle: LUER LOCK type. Sterility & Product and packaging: Sterile, ethylene oxide sterilisation and compiles to BS EN 550:1994-ISO 11135-1:2007 or its equivalent Indian Standards) Primary packaging: Each syringe packed in an individual sterilized peel-pack / blister pack / ribbon pack. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	200000	piece
1.034	<p>Sterile Disposable syringe for single use without needle 50 ml Luer Lock Type: Specifications: Syringes should be with clear syringe barrel, precise graduation in ml with permanent marking, easy readability, Safe plunger backstop. Syringe capacity: 50 ml Syringe Material: medical grade plastic (Polyethylene (PE)/ polypropylene (PP)/ polystyrene (PS)). The materials used for fabrication of syringes (including lubricant) and packaging should be non toxic, do not adversely affect the efficacy, safety and acceptability. Syringe Barrel: sufficiently transparent to allow easy measurement of the volume contained in the syringe and detection of air bubbles. Length with a maximum usable capacity of at least 10% more than the nominal capacity. Standards: Indian standard IS 10258 -2002 latest, equivalent to ISO 7886 part 3. Barrel Nozzle: LUER LOCK type. Sterility & Product and packaging: Sterile, ethylene oxide sterilisation and compiles to BS EN 550:1994-ISO 11135-1:2007 or its equivalent Indian Standards) Primary packaging: Each syringe packed in an individual sterilized peel-pack / blister pack / ribbon pack. Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	150000	piece
1.035	<p>ABG syringes 1 ml capacity: A balanced heparin containing ABG syringe with needle for blood gas analysis. Complete data sheet to be provided</p>	350000	piece

1.036	<p>Sterile disposable hypodermic needles specifications: 16G x 1.5" Needle material: Stainless steel tubing conform to IS 10654:2002 latest for sterile hypodermic needles, equivalent to ISO 9626 and needle point in accordance with ISO 7864. Should be packed without any risk of breach in sterility. ETO/Gamma Sterilized. Needle Measurements: External diameter expressed in Gauge and inches. Needle Protective cap: Medical grade plastic. Dead space: Should be minimum to reduce waste. Sterile: By Ethylene oxide and compiles to BS EN 550:1994-ISO 11135-1:2007. Packaging and labelling: Primary packaging: Each needle packed in an individual sterilized peel-pack made of paper and/or plastic. The materials used for fabrication of needles and packaging do not adversely affect the efficacy, safety and acceptability Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	400000	piece
1.037	<p>Sterile disposable hypodermic needles specifications: 18G x 1.5" Needle material: Stainless steel tubing conform to IS 10654:2002 latest for sterile hypodermic needles, equivalent to ISO 9626 and needle point in accordance with ISO 7864. Should be packed without any risk of breach in sterility. ETO/Gamma Sterilized. Needle Measurements: External diameter expressed in Gauge and inches. Needle Protective cap: Medical grade plastic. Dead space: Should be minimum to reduce waste. Sterile: By Ethylene oxide and compiles to BS EN 550:1994-ISO 11135-1:2007. Packaging and labelling: Primary packaging: Each needle packed in an individual sterilized peel-pack made of paper and/or plastic. The materials used for fabrication of needles and packaging do not adversely affect the efficacy, safety and acceptability Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	800000	piece
1.038	<p>Sterile disposable hypodermic needles specifications: 20G x 1.5" Needle material: Stainless steel tubing conform to IS 10654:2002 latest for sterile hypodermic needles, equivalent to ISO 9626 and needle point in accordance with ISO 7864. Should be packed without any risk of breach in sterility. ETO/Gamma Sterilized. Needle Measurements: External diameter expressed in Gauge and inches. Needle Protective cap: Medical grade plastic. Dead space: Should be minimum to reduce waste. Sterile: By Ethylene oxide and compiles to BS EN 550:1994-ISO 11135-1:2007. Packaging and labelling: Primary packaging: Each needle packed in an individual sterilized peel-pack made of paper and/or plastic. The materials used for fabrication of needles and packaging do not adversely affect the efficacy, safety and acceptability Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	200000	piece
1.039	<p>Sterile disposable hypodermic needles specifications: 21G x 1.5" Needle material: Stainless steel tubing conform to IS 10654:2002 latest for sterile hypodermic needles, equivalent to ISO 9626 and needle point in accordance with ISO 7864. Should be packed without any risk of breach in sterility. ETO/Gamma Sterilized. Needle Measurements: External diameter expressed in Gauge and inches. Needle Protective cap: Medical grade plastic. Dead space: Should be minimum to reduce waste. Sterile: By Ethylene oxide and compiles to BS EN 550:1994-ISO 11135-1:2007. Packaging and labelling: Primary packaging: Each needle packed in an individual sterilized peel-pack made of paper and/or plastic. The materials used for fabrication of needles and packaging do not adversely affect the efficacy, safety and acceptability Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	200000	piece

1.040	<p>Sterile disposable hypodermic needles specifications: 22G x 1.5" Needle material: Stainless steel tubing conform to IS 10654:2002 latest for sterile hypodermic needles, equivalent to ISO 9626 and needle point in accordance with ISO 7864. Should be packed without any risk of breach in sterility. ETO/Gamma Sterilized.</p> <p>Needle Measurements: External diameter expressed in Gauge and inches. Needle Protective cap: Medical grade plastic. Dead space: Should be minimum to reduce waste. Sterile: By Ethylene oxide and compiles to BS EN 550:1994-ISO 11135-1:2007. Packaging and labelling: Primary packaging: Each needle packed in an individual sterilized peel-pack made of paper and/or plastic. The materials used for fabrication of needles and packaging do not adversely affect the efficacy, safety and acceptability Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	150000	piece
1.041	<p>Sterile disposable hypodermic needles specifications: 23G x 1.0" Needle material: Stainless steel tubing conform to IS 10654:2002 latest for sterile hypodermic needles, equivalent to ISO 9626 and needle point in accordance with ISO 7864. Should be packed without any risk of breach in sterility. ETO/Gamma Sterilized.</p> <p>Needle Measurements: External diameter expressed in Gauge and inches. Needle Protective cap: Medical grade plastic. Dead space: Should be minimum to reduce waste. Sterile: By Ethylene oxide and compiles to BS EN 550:1994-ISO 11135-1:2007. Packaging and labelling: Primary packaging: Each needle packed in an individual sterilized peel-pack made of paper and/or plastic. The materials used for fabrication of needles and packaging do not adversely affect the efficacy, safety and acceptability Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	900000	piece
1.042	<p>Sterile disposable hypodermic needles specifications: 24G x 1.0" Needle material: Stainless steel tubing conform to IS 10654:2002 latest for sterile hypodermic needles, equivalent to ISO 9626 and needle point in accordance with ISO 7864. Should be packed without any risk of breach in sterility. ETO/Gamma Sterilized.</p> <p>Needle Measurements: External diameter expressed in Gauge and inches. Needle Protective cap: Medical grade plastic. Dead space: Should be minimum to reduce waste. Sterile: By Ethylene oxide and compiles to BS EN 550:1994-ISO 11135-1:2007. Packaging and labelling: Primary packaging: Each needle packed in an individual sterilized peel-pack made of paper and/or plastic. The materials used for fabrication of needles and packaging do not adversely affect the efficacy, safety and acceptability Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	400000	piece
1.043	<p>Sterile disposable hypodermic needles specifications: 26G x 0.5" Needle material: Stainless steel tubing conform to IS 10654:2002 latest for sterile hypodermic needles, equivalent to ISO 9626 and needle point in accordance with ISO 7864. Should be packed without any risk of breach in sterility. ETO/Gamma Sterilized.</p> <p>Needle Measurements: External diameter expressed in Gauge and inches. Needle Protective cap: Medical grade plastic. Dead space: Should be minimum to reduce waste. Sterile: By Ethylene oxide and compiles to BS EN 550:1994-ISO 11135-1:2007. Packaging and labelling: Primary packaging: Each needle packed in an individual sterilized peel-pack made of paper and/or plastic. The materials used for fabrication of needles and packaging do not adversely affect the efficacy, safety and acceptability Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	1200000	piece

1.044	<p>Sterile disposable hypodermic needles specifications: 31G x 5/16" (8 mm) Needle material: Stainless steel tubing conform to IS 10654:2002 latest for sterile hypodermic needles, equivalent to ISO 9626 and needle point in accordance with ISO 7864. Should be packed without any risk of breach in sterility. ETO/Gamma Sterilized. Needle Measurements: External diameter expressed in Gauge and inches. Needle Protective cap: Medical grade plastic. Dead space: Should be minimum to reduce waste. Sterile: By Ethylene oxide and complies to BS EN 550:1994-ISO 11135-1:2007. Packaging and labelling: Primary packaging: Each needle packed in an individual sterilized peel-pack made of paper and/or plastic. The materials used for fabrication of needles and packaging do not adversely affect the efficacy, safety and acceptability Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility. or certificates to be available in the OEM website .</p>	1000000	piece
1.045	<p>Infusion (IV) Set, non-vented: Sterile, disposable, single use, gravity feed, non-vented infusion set confirming to IS: 12655 (Part -4) latest. IV set should be non-toxic, pyrogen free, non-kinking quality. IV set should have sturdy spike with protective cap, drip chamber should have fluid filter, sturdy roller clamp, non-kinking, non-toxic tubing, stepwise injection port, luer slip connector with 21 G needle. Infusion should be macro drip (20 drops = 1 ml ± 0.1ml) Pack without any risk of breach in sterility, ETO/ Gamma sterilized. Outer cover should have option eye cut/tear mark/peel-pack for easy opening. ISO or ISI certified or its equivalent Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility or certificates/documents to be available in the OEM website .</p>	400000	piece
1.046	<p>Infusion (IV) Set, vented: Sterile, disposable, single use, gravity feed, vented infusion set confirming to IS: 12655 (Part -4) latest. IV set should be non-toxic, pyrogen free, non-kinking quality. IV set should have sturdy spike with protective cap, drip chamber should have fluid filter and non-leaking air vent with provision of closure of airvent, sturdy roller clamp, non-kinking, non-toxic tubing, stepwise injection port, luer slip connector with 21 G needle. Infusion should be macro drip (20 drops = 1 ml ± 0.1ml) Pack without any risk of breach in sterility, ETO/ Gamma sterilized. Outer cover should have option eye cut/tear mark/peel-pack for easy opening. ISO or ISI certified or its equivalent Certificates: OEM should produce Data sheet, certificate of standards, valid licenses, manufacturing units certificates (ISO 13485), Marketing approval / test reports / proof of sterility or certificates/documents to be available in the OEM website .</p>	400000	piece
1.047	<p>3 ply Surgical Mask Tie type: Specifications: 1) ISO certified: yes 2) Fabric material: Non-woven fabric with middle polyfibre layer 3) GSM of the fabric(G/m2): 60-70 4) Pleated construction of Mask: to provide adjustable and leakproof fitting 5) No. of layers of plies of fabric: 3 layers 6) Bacterial filtration efficiency (BFE%): >=99 7) Splash resistant: Yes 8) Breathing resistance (mm of H2O/cm2): <=5(splash resistant) 9) Securing type: Tie type 10) Adjustable flexible nose clip: with 11) Colour: any 12) Length : 17 - 18cm 13) Width: 9 - 10cm 14) Submission of Test Report to the buyer on demand from NABL/ILAC accredited or Central Govt. Lab to prove conformity of products to the specification: Yes 15) Certification: ISI</p>	1500000	piece
1.048	<p>3 ply Surgical Mask Earloop type: Specifications: 1) ISO certified: yes 2) Fabric material: Non-woven fabric with middle polyfibre layer 3) GSM of the fabric(G/m2): 60-70 4) Pleated construction of Mask: to provide adjustable and leakproof fitting 5) No. of layers of plies of fabric: 3 layers 6) Bacterial filtration efficiency (BFE%): >=99 7) Splash resistant: Yes 8) Breathing resistance (mm of H2O/cm2): <=5(splash resistant) 9) Securing type: Earloop type 10) Adjustable flexible nose clip: with 11) Colour: any 12) Length : 17 - 18cm 13) Width: 9 - 10cm 14) Submission of Test Report to the buyer on demand from NABL/ILAC accredited or Central Govt. Lab to prove conformity of products to the specification: Yes 15) Certification: ISI</p>	1000000	piece
1.049	<p>Doctors Cap (Bouffant type) Specifications:- 1) Size: 18-21 inch 2) Type of material: Non-woven breathable 3) Design: Bouffant type 4) Composition of nonwoven material: spun-bond non-woven 5) GSM of Cap material: 20gm/sq.mtrs. 6) Well fitted with double elastic or string: Yes 7) Properly ultrasonically sealed for uniformity: Yes 8) Colour: Any 9) Bacterial filtration: Yes 10) Air permeable: Yes 11) Certification: ISO certified 12) Submission of Test Report to the buyer from NABL/ILAC accredited or Central Govt. lab to prove conformity of products to the specification on demand to the buyer: Yes</p>	100000	piece

1.050	Surgeon's Cap with sweat band : Specifications:- 1) Size: 18-21 inch 2) Type of material: Non-woven breathable 3) Composition of nonwoven material: spun-bond non-woven 4) Well fitted with double elastic or string: Yes 5) Properly ultrasonically sealed for uniformity: Yes 6) Colour: Any 7) Bacterial filtration: Yes 8) Air permeable: Yes	300000	piece
1.051	Sterile Medical Gown Specifications:- 1) Gown type: ICU gown 2) Colour: Medical blue 3) Disposable: Yes 4) Breathable: Yes 5) antistatic properties: yes 6) Impervious to fluids: Yes 7) Reinforced: No 8) Sterilized: Yes 9) Method of sterilization: ETO 10) Gown width: 70-80cm 11) Gown length: 125-150cm 12) Adjustable neck ties : Yes. 13) No. of belts to tie: 2 pairs. 14) material used for gown: Non-woven, spun bond 15) GSM of non-woven fabric: 40 g/m2 16) Alcohol repellent material: Yes 17) Sleeve type: Full sleeve 18) Cuff end: With knitted cuff at least 10cm width 19) Wrap around type: Yes 20) Type of packing: Sterile pack, pouch must be peel off made of medical grade paper 21) AAMI protection level: Level III 22) Certification : ISO certified 23) submission of test report clearly specifying the nomenclature of fabric, method of sterilization and AAMI protection level in addition to other parameter to the buyer on form 39 or from NABL/ILAC accredited or Central Govt. lab to prove conformity certificate to be furnished to buyer at the time of supply: Yes	100000	piece
1.052	Unsterile Medical Gown Specifications:- 1) Gown type: Ward gown 2) Colour: Medical blue 3) Disposable: Yes 4) Breathable: Yes 5) antistatic properties: yes 6) Impervious to fluids: Yes 7) Reinforced: No 8) Sterilized: No 10) Gown width: 70-80cm 11) Gown length: 125-150cm 12) Adjustable neck ties : Yes. 13) No. of belts to tie: 2 pairs. 15) material used for gown: Non-woven, spun bond 16) GSM of non-woven fabric: 40 g/m2 17) Alcohol repellent material: Yes 18) Sleeve type: Full sleeve 19) Cuff end: With knitted cuff at least 10cm width 20) Wrap around type: Yes.	100000	piece
1.053	Surgeon Gown for HIV protection Specifications:- 1) Gown type: surgeon gown for HIV protection 2) Colour: Medical blue / green 3) Disposable: Yes 4) Breathable: Yes 5) Antistatic properties: yes 6) Impervious to fluids: Yes 7) Reinforced: yes 8) Sterilized: Yes 9) Method of sterilization: ETO 10) Gown width: 70-80cm 11) Gown length: 125-150cm . 12) No. of belts to tie: 2 pairs with paper clip 13) Material used for gown: Non woven SMS 14) Composition of reinforced material : SMS+Poly coated to be smooth inner surface 15) Reinforced critical zones (front and sleeve up to elbows) to provide 100% protection against any amount of fluid: Yes 16) GSM of non woven fabric: 68g/m2 and above 17) Closing at neck: Hook and loop / velcro fastener with piping of non woven material stitched to raw edges of neck opening extending at both ends for tying 18) Pair of disposable absorbent hand wipes: Yes 19) Sleeve type: Raglan sleeve, full sleeve 20) Cuff end: With knitted cuff of at least 10cm width 21) Wrap around type: Yes 22) Type of packing : Sterile pack, must be double packed with outer pouch must be peel off made of medical grade paper 23) type wrapping: Set in sleeve 24) AAMI Protection level: Level IV 25) Certification: ISO 13485:2016 26) Submission of test report clearly specifying the nomenclature of fabric, method of sterilization an AAMI protection level in addition to other parameter to the buyer on form 39 or from NABL/ILAC accredited or Central Govt. Lab to prove conformity of products to the specification : yes 27) Product conformity certificate to be furnished to buyer at the time of supply: Yes	120000	piece
1.054	SMS Sterilization wrapping paper sheet (blue/green), Size: 100 x 100 cms Specifications: Non-woven, bacterial barrier, single use, light weight, wet penetration resistant. Lowering the risk of puncturing or tearing during wrapping, transport. Comply with EN 868 – 2. Comply with the essential requirement 93/42/EEC Directive and should be CE marked or equivalent. Can be sterilized in steam, EO, Radiation, Formaldehyde and VH2 O2 .	150000	piece
1.055	SMS Sterilization wrapping paper sheet (blue/green), Size: 150 x 150 cms Specifications: Non-woven, bacterial barrier, single use, light weight, wet penetration resistant. Lowering the risk of puncturing or tearing during wrapping, transport. Comply with EN 868 – 2. Comply with the essential requirement 93/42/EEC Directive and should be CE marked or equivalent. Can be sterilized in steam, EO, Radiation, Formaldehyde and VH2 O2 .	150000	piece
1.056	Disposable patient leggings Specifications:- knee length size, Polycoated, laminated, non-woven with tiable hem, impervious, stretchable	40000	piece

1.057	Bootleg leggings Specifications:- 1) Colour: any 2) Skid resistance: Yes 3) Dust proof: Yes 4) Well stitched in universal regular size: Yes 5) Should cover the knee: Yes. length should be above knee level. 6) hard elasticized for better grip and easy wear: Yes. Should provide one pair of ties additionally. 7) Disposable: yes 8) Sterile: No 9) Packing: Pair individually packed 10) Shoe size for which item is suitable: 10, 11) Clear indication of size on disposable shoe cover: Yes 12) Material (Non-toxic, medical grade): Non-woven thick fabric. Impervious to fluids:Yes 13) Composition of non-woven material: Spun bond + non woven 14) GSM of non woven material: 40 g/sq.mtrs 16) Submission of Test Report to the buyer on demand from NABL/ILAC accredited or Central Govt. Lab to prove conformity of products to the specification on demand to the buyer: Yes	50000	piece
1.058	Medical staff Apron Specifications:- 1)single use , sterile, disposable medical staff apron with proper neck and arm hole cutting with straps with good resistance to tear 2) Material of apron: Medical grade PE 3) GSM of apron material: > 40 gms per sq.mtrs 4) Thickness: 50 micron and above 5) Width of apron : 65- 75 cms 6) Length of apron: 120 - 140 cms 7)Colour: Blue / green . 8)Sterile: Yes 9) Packing: Individually packed in sterile manner. Submission of test report to the buyer on Form 39 or NaBL/ILAC accredited or Central Govt. Lab to prove conformity of products to the specification on demand: Yes	1000000	piece
1.059	Surgical Drapes (OT Sheet) Specifications: Sterile disposable surgical drapes (OT sheet), non-woven material - 100 x 100 cm with opening of 14 x 18 cm in the centre with adhesive borders.	200000	piece
1.060	Microporus surgical Paper Tape adhesive type : Size:- 2.5 cm x 9 m, Air permeable, hypo allergic with good skin adhesion.	70000	piece
1.061	Disposable umbilical cord clamp Specifications: Single use packed Sterile, ETO sterilized. Polyoxymethylene copolymer/plastic material Interlocking teeth for security lock with a click reopening or slipping when clamped Smooth Edges 5-6 cm long, width 1-1.5 cm when closed	20000	piece
1.062	Self adhesive Sterilization Indicator Tapes for Steam Sterilizer: Material of tape: Crepe / semi crepeed paper. Tape Length: 90 meters. Tape width: 18 mm. Shelf life: 2 years. Colour of tape before steriliation : White. Change of the integrated chemical indicator on the tape by its display of diagonal stripes: yes After the sterilization process is complete, the tape can easily be removed with no adhesive residue:yes.	2200	piece
1.063	Wrapper for ETO Sterilization :Size : 7.5cm x 200 mtr. Medical grade sheet should be made of high density polyethylene(HDPE) covers for ETO Sterilization. Wrapper should allows for an effective sterilant concentration to be achieved with inbuilt indicator with resistance to microbial penetration, good tear strength, puncture resistance and clean peel. Inbuilt indicator: Both EO and steam indicators should be printed on the sides. Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light. Should comply to standards EN ISO 11140-1 , EN ISO 868-3,5 and 6. Paper and plastic as per EN ISO standards. Leakage test method→ ASTM F-1929-98. Bioburden EN ISO 11607. A medical grade 60gsm paper confirming EN 868-3. Brochure to be provided.	200	roll
1.064	Wrapper for ETO Sterilization :Size : 15cm x 200 mtr. Medical grade sheet should be made of high density polyethylene(HDPE) covers for ETO Sterilization. Wrapper should allows for an effective sterilant concentration to be achieved with inbuilt indicator with resistance to microbial penetration, good tear strength, puncture resistance and clean peel. Inbuilt indicator: Both EO and steam indicators should be printed on the sides. Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light. Should comply to standards EN ISO 11140-1 , EN ISO 868-3,5 and 6. Paper and plastic as per EN ISO standards. Leakage test method→ ASTM F-1929-98. Bioburden EN ISO 11607. A medical grade 60gsm paper confirming EN 868-3. Brochure to be provided.	300	roll

1.065	<p>Wrapper for ETO Sterilization :Size: 20cm x 200 mtr. Medical grade sheet should be made of high density polyethylene(HDPE) covers for ETO Sterilization. Wrapper should allows for an effective sterilitant concentration to be achieved with inbuilt indicator with resistance to microbial penetration, good tear strength, puncture resistance and clean peel. Inbuilt indicator: Both EO and steam indicators should be printed on the sides. Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light.</p> <p>Should comply to standards EN ISO 11140-1 , EN ISO 868-3,5 and 6. Paper and plastic as per EN ISO standards. Leakage test method→ ASTM F-1929-98. Bioburden EN ISO 11607.</p> <p>A medical grade 60gsm paper confirming EN 868-3. Brochure to be provided.</p>	200	roll
1.066	<p>Wrapper for ETO Sterilization :Size : 30cm x 200 mtr. Medical grade sheet should be made of high density polyethylene(HDPE) covers for ETO Sterilization. Wrapper should allows for an effective sterilitant concentration to be achieved with inbuilt indicator with resistance to microbial penetration, good tear strength, puncture resistance and clean peel. Inbuilt indicator: Both EO and steam indicators should be printed on the sides. Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light.</p> <p>Should comply to standards EN ISO 11140-1 , EN ISO 868-3,5 and 6. Paper and plastic as per EN ISO standards. Leakage test method→ ASTM F-1929-98. Bioburden EN ISO 11607.</p> <p>A medical grade 60gsm paper confirming EN 868-3. Brochure to be provided.</p>	200	roll
1.067	<p>Wrapper for Plasma Sterilization (Low temperature hydrogen peroxide Plasma Sterilizer): Size : 3" x 70 mtr. Medical grade sheet, should be made of high density polyethylene (HDPE) and allows for an effective sterilitant concentration to be achieved in low temperature hydrogen peroxide gas sterilization to sterilization of heat liable devices. wrapper should be with inbuilt indicator with resistance to microbial penetration, good tear strength, puncture resistance and clean peel. Inbuilt indicator: Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light. Comply with EN 868-5 and ISO 11607. Free from lead and heavy metals. Brochure to be provided.</p>	200	roll
1.068	<p>Wrapper for Plasma Sterilization (Low temperature hydrogen peroxide Plasma Sterilizer): Size : 6" x 70 mtr. Medical grade sheet, should be made of high density polyethylene (HDPE) and allows for an effective sterilitant concentration to be achieved in low temperature hydrogen peroxide gas sterilization to sterilization of heat liable devices. wrapper should be with inbuilt indicator with resistance to microbial penetration, good tear strength, puncture resistance and clean peel. Inbuilt indicator: Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light. Comply with EN 868-5 and ISO 11607. Free from lead and heavy metals. Brochure to be provided.</p>	200	roll
1.069	<p>Wrapper for Plasma Sterilization (Low temperature hydrogen peroxide Plasma Sterilizer) Size : 10" x70 mtr. Medical grade sheet, should be made of high density polyethylene (HDPE) and allows for an effective sterilitant concentration to be achieved in low temperature hydrogen peroxide gas sterilization to sterilization of heat liable devices. wrapper should be with inbuilt indicator with resistance to microbial penetration, good tear strength, puncture resistance and clean peel. Inbuilt indicator: Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light. Comply with EN 868-5 and ISO 11607. Free from lead and heavy metals. Brochure to be provided.</p>	200	roll
1.070	<p>Wrapper for Plasma Sterilization (Low temperature hydrogen peroxide Plasma Sterilizer): Size : 20" x 70 mtr. Medical grade sheet, should be made of high density polyethylene (HDPE) and allows for an effective sterilitant concentration to be achieved in low temperature hydrogen peroxide gas sterilization to sterilization of heat liable devices. wrapper should be with inbuilt indicator with resistance to microbial penetration, good tear strength, puncture resistance and clean peel. Inbuilt indicator: Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light. Comply with EN 868-5 and ISO 11607. Free from lead and heavy metals. Brochure to be provided.</p>	50	roll
1.071	<p>ETO Cartridges : Cartridge containing 200gm of 100% ethylene oxide gas, for use in Ethylene Oxide Sterilizer (Capacity : 452 litres). ISO Certified. Brochure to be provided.</p>	400	cartridge
1.072	<p>Rust Removing soak solution: non toxic, biodegradable, water soluble, non corrosive, non flammable. Capacity 1 litre</p>	30	bottle

1.073	<p>Rapid Multi Enzyme cleaner for manual washing, 1 litre capacity :</p> <ul style="list-style-type: none"> • Multiple enzymatic concentrate solution with lipases, cellulases, proteases and amylases • Must act against all possible body secretions, compatible to all instruments • Fast acting within 2- 10 minutes, allow safe disposable via the drain after use • Leaves no residue after wash, non corrosive, nonpungent odour, Stable formulation with near neutral pH • Compliance to ISO 13485 	100	bottle
1.074	<p>Ultra Rapid Multi Enzyme cleaner for Automatic and washing disinfectors, 1 litre capacity:</p> <ul style="list-style-type: none"> • Multiple enzymatic concentrate solution with lipases, cellulases, proteases and amylases • Film based Polymer technology with controlled enzyme release • Must act against all possible body secretions, compatible to all instruments • Fast acting within 2- 10 minutes, allow safe disposable via the drain after use • Leaves no residue after wash, non corrosive, nonpungent odour, Stable formulation with near neutral pH • Should have scale inhibition properties • Should help retain shine of the instrument • Should be usable for manual, ultra sonic and automatic cleaning systems • Compliance to ISO 13485 	200	bottle
1.075	<p>Instrument cleaning Brushes Specifications: Tooth brush style, plastic handle , 7-10 inches long, polypropylene bristles, single ended, three rows</p>	50	piece
1.076	<p>General Instrument cleaning Brushes Specifications: around 7-10 inches long, Stainless Steel Brushes</p>	50	piece
1.077	<p>Bowie-Dick steam process indicator Specifications:- disposable single use pack, lead and heavy metal free, ability to preserve indicator sheet for permanent record. Stable colour, colour change is not reversible, indicator ink should not fade in storage or under fluorescent light. comparative images for various results and conclusions should be provided in the pack. Add each pack to have lot numbers and expiry date and a uniform contrast colour change. Brochure to be provided</p>	300	piece
1.078	<p>Biological Indicator for Steam Sterilization Specifications:- self contained Geobacillus Stearothermophilus - ISO 11138:2017 Parts 1 and 3, ISO 11138-7:2019(en) and european standard certified Biological test pack with auto readers for steam sterilizer, smart read, Stable colour, colour change should not to be reversible, indicator ink should not fade in storage or under fluorescent light. Spore population in the Indicator: More than 10^5 CFU . Time taken for Biological Indicator Result: rapid read out (within 30 min preferably). Readout technology: Initially by fluorescence detection and later by PH colour change. Biological indicator ampule should be integrated with type 5 chemical indicators additionally. Shelf life: 2 years. Quality certification report regarding Spore Population,Z Value,D Value, to be provided. brochure for the BI and incubator to be provided or should be available in the website. compatible incubator should be provided free of cost</p>	3000	vial
1.079	<p>Biological indicator for ETO sterilization Specifications: self-contained biological indicator (BI) that consists of a Bacillus atrophaeus, ISO 11138-1:2017 and ISO 11138-2:2017 standards for ETO sterilizer. rapid read out , smart read, Stable colour, colour change should not to be reversible, indicator ink should not fade in storage or under fluorescent light. Shelf life: 2 years. Time taken for Biological Indicator Result: rapid read out (within 4 hours preferably). Readout technology: Initially by fluorescence detection and later by PH colour change. Spore population in the Indicator: More than 10^6CFU . Readout technology: Initially by fluorescence detection later by PH colour change. Biological indicator ampule should be integrated with type 4 / type 5 chemical indicators additionally. Quality certification report regarding Spore Population,Z Value,D Value, to be provided. brochure for the BI and incubator to be provided or should be available in the website. compatible incubator should be provided free of cost</p>	700	vial

1.080	Biological Indicator for Plasma Sterilizer specifications:- self contained Geobacillus Stearothermophilus - ISO 11138:2017 Parts 1 and 3, ISO 11138-7:2019(en) and european standard certified Biological test pack with auto readers for plasma sterilizer, smart read, Stable colour, colour change should not to be reversible, indicator ink should not fade in storage or under fluourescent light. Spore population in the Indicator: More than 10 ⁵ CFU . Time taken for Biological Indicator Result: rapid read out (within 30 min preferably). Readout technology: Initially by fluorescence, later by PH colour change. Biological indicator ampule should be integrated with type 4 / type 5 chemical indicators additionally. Shelf life: 2 years.Quality certification report regarding Spore Population,Z Value,D Value, to be provided. brochure for the BI and incubator to be provided or should be available in the website. compatible incubator should be provided free of cost	1000	vial
1.081	Steam Process Indicator Strip EN ISO-11140-1 : 2005 certified Class V Steam Sterilizer. Single use strips for steam sterilizer. compatible with 121°C at 20 minutes/ 134°C at 3 minutes process. Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light. It should contain colour comparison option after sterilisation for easy confirmation. Indicator color should not change in dry heat. Strip with adhesive option will be preferable for easy documentation. Brochure to be provided	150000	strip
1.082	ETO Process Indicator Strip Specification:-Process indicator should be according to EN ISO 11140 – 1, Class 4 / class 5 indicator for EO gas sterilization. Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light. It should contain colour comparison option after sterilisation for easy confirmation.Should be lead free. Indicator strip with adhesive option will be preferable for easy documentation. Brochure to be provided	3000	strip
1.083	Plasma (H2O2) Process Indicator Strip Specification:-Process indicator should be according to EN ISO 11140 – 1,Class 4 / class 5 indicator for H2O2 plasma sterilization. Change of colour after sterilisation should be contrast with initial color and colour change should not be reversible. Indicator ink should not fade in storage or under fluourescent light. It should contain colour comparison option after sterilisation for easy confirmation.Should be lead free. Indicator strip with adhesive option will be preferable for easy documentation. Brochure to be provided	3000	strip
1.084	Batch Monitoring chemical Indicator strip for Bowie Dick Test : Air Removal Chemical Indicator strip (Type 2) for Bowie Dick process challenge device (PCD) with adhesive option for easy documentation. Indicator color should not change in dry heat. Air Removal Test according to EN 285 (7kg) or AAMI/ANSI/ST79 (4kg Test) and Validated according to the test method of ISO 11140-1+5 for steam sterilization. 10 nos. of Stainless steel PCD - EU version, according to EN ISO 11140-4 and hollow load test according to EN 867-5 (Brochure to be provided) to be provided free of cost. Complete Data sheet of both batch monitoring chemical Indicator for Bowie Dick test/ PCD and sample should be submitted. 10 Nos. of PCD device should be provided on free of cost.	3000	strip
1.085	Batch Monitoring Chemical Indicator strip for steam sterilization: Type 2 chemical indicator to be used in Compact PCD for steam sterilization according to EN ISO - 11140-1 and lab report EN ISO 17025. Indicators should have adhesive for easy documentation and should not change color in dry heat as per EN ISO - 11140. 10 Nos. of compatible Compact stainless steel Process Challenging Device (PCD) – Specific Load Test according to EN ISO 11140-1.(Brochure to be provided) should be provided on free of cost. Complete Data sheet of both batch monitoring chemical Indicator / PCD and sample should be submitted. 10 Nos. of PCD device should be provided on free of cost.	15000	strip
1.086	Batch Monitoring Chemical Indicator strip for EO sterilization : Class-2 Chemical indicator to monitor ethylene oxide sterilization compatible to the stainless-steel helix PCD as per EN 1422. Should be lead free. Should be with adhesive for easy documentation. 2 Nos. of compatible Metal stainless steel Process Challenge Device (PCD) for Batch Monitoring System (BMS) -according to EN 1422 Type test for large sterilizer as per EN 285 (Brochure to be provided) should be provided on free of cost. Complete Data sheet of both batch monitoring chemical Indicator / PCD and sample should be submitted. 2 Nos. of PCD device should be provided on free of cost.	800	strip

1.087	Batch Monitoring Chemical Indicator strip for H2O2 (Plasma) sterilization Type 2 chemical indicator to be used in Helix PCD according to EN ISO - 11140-1 and lab report EN ISO 17025. Indicators should have adhesive for easy documentation. 3 Nos. of compatible Helix Process Challenge Device (PCD) for Batch Monitoring System (BMS) according to EN ISO 11140-1 (Brochure to be provided) should be provided on free of cost. Complete Data sheet of both batch monitoring chemical Indicator / PCD and sample should be submitted. 3 Nos. of PCD device should be provided on free of cost.	3000	strip
1.088	Documentation Label for Steam Sterilization Specification: - 3 line Documentation label with integrated steam indicator with adhesive base. Should comply with EN ISO 11140-1. Indicator colour should change only in steam sterilization, not in dry heat. Rate must be Quoted per label, not per roll. 3 Nos. Label guns with ink should be provided for each 25000 Nos. of labels, on loan basis.	75000	strip
1.089	Documentation Label for ETO Sterilization. Specification: - 3 line Documentation label with integrated steam indicator with adhesive base. Should comply with EN ISO 11140-1. Rate must be Quoted per label, not per roll. 2Nos. of Label gun with ink should be provided for each 25000 Nos. of labels, on loan basis.	25000	strip
1.090	Documentation Label for Plasma Sterilization. Specification: - 3 line Documentation label with integrated steam indicator with adhesive base. Should comply with EN ISO 11140-1. Rate must be Quoted per label, not per roll. 2 Nos. of Label gun with ink should be provided for each 25000 Nos. of labels, on loan basis.	60000	strip
1.091	Spinal Needle with Whitacre Pencil Point Size: 25G x 90mm Specifications: - Sterile, disposable, Whitacre pencil point, individually packed. Graduated metal introducer & translucent hub with colour coding, should be of good quality material and pack without any risk of breach in sterility. ETO/Gamma sterilized.	5000	piece
1.092	Spinal Needle with Quincke Type Point Size: 23G x 90mm Specifications: - Sterile, disposable, Quincke type point, individually packed. Translucent hub with colour coding, should be of good quality material and pack without any risk of breach in sterility. ETO/Gamma sterilized.	10000	piece
1.093	Spinal Needle with Quincke Type Point Size: 25G x 90mm Specifications: - Sterile, disposable, Quincke point, individually packed. Translucent hub with colour coding, should be of good quality material and pack without any risk of breach in sterility. ETO/Gamma sterilized.	10000	piece
1.094	Colour Coding Tag for surgical instruments: Indented to adhere to surgical instruments for identification, organization, management using various colours. Should be able to withstand repeated sterilization cycles suitable for 121 degree C and 134 degree C. Price should be quoted per tag not for sheet (Brochure to be attached)	10000	piece
1.095	Sterilization bins for steam sterilization: Sterilization bins cylindrical type, medium size measuring 12" height 14" diameter with provision of holes for steam entry and manual locking systems.	50	piece
1.096	Heat Resistant Gloves: Specifications: Long (upto elbow), resistant to high temperature, to be used in Autoclave, heat resistant and preserved material, smooth finish with good hand grip and full fingered, large size (24 inches)	50	pair
1.097	Thermal Printer Roll: compatible for use in Ethylene Oxide Sterilizer (capacity : 452 litres). Size : 2.3" x 10 mtr. Roll	25	roll
1.098	Heating Element, 6Kw capacity, compatible for use in microprocessor controlled, high pressure, fully automatic, 36KW, double door Steam Sterilizer.	50	piece
1.099	Heating Element Washer, compatible for use in microprocessor controlled, high pressure, fully automatic, 36KW, double door Steam Sterilizer, functioning in CSSD, JIPMER.	50	piece
1.100	Thermo seal Tape: Used to pack the surgical trays for steam sterilization. size 18mm x 50mts. Material: Crepe paper with natural rubber adhesive for medical use. Should sustain temperature of 134°C in a autoclave. Should not leave the adhesive behind after sterilization.	1200	roll
1.101	Sterile Cotton Gauze with radio-opaque line marker, Size: 5 x 5 cms, 16ply, type 17 The marker should be made up of multifilament yarn with at least 60% barium coated which will not break or come out separately during usage. Each should have folded edges to avoid fraying of the fabric. The material should have been bleached using hydrogen peroxide. Each pack should have ETO sterilized with double sealed blister pack for secured sterility with sterility indicator.	50000	piece

1.102	Sterile Cotton Gauze with radio-opaque line marker, Size: 7.5 x 7.5 cms, 16ply, type 17 The marker should be made up of multifilament yarn with at least 60% barium coated which will not break or come out separately during usage. Each should have folded edges to avoid fraying of the fabric. The material should have been bleached using hydrogen peroxide. Each pack should have ETO sterilized with double sealed blister pack for secured sterility with sterility indicator.	50000	piece
1.103	Sterile Cotton Gauze with radio-opaque line marker, Size: 10 x 10 cms, 8ply, type 17 The marker should be made up of multifilament yarn with at least 60% barium coated which will not break or come out separately during usage. Each should have folded edges to avoid fraying of the fabric. The material should have been bleached using hydrogen peroxide. Each pack should have ETO sterilized with double sealed blister pack for secured sterility with sterility indicator.	25000	piece
1.104	Sterile Surgical Mopping Pad with radio-opaque line marker and a coloured loop, Size: 30 x 30 cms, 8ply, type 17. The marker should be made up of multifilament yarn with at least 60% barium coated which will not break or come out separately during usage. Each should have folded and stitched edges to avoid fraying of the fabric. The material should have been bleached using hydrogen peroxide. Each pack should have ETO sterilized with double sealed blister pack for secured sterility with sterility indicator. There should be a coloured loop for identification.	20000	piece
1.105	Sterilization containers with compatible SS basket for steam sterilization : Dimesion of container in mm: 1 STU (600 X 300 X 260), inner basket measuring (540x240x180): stainless steel container should be AISI 304 grade, an austentic stainless steel with exceptional mechanical resistance and highly resistant corrosion. Container lid should have to be provided with a protected valve system that assures a mechanical control of steam penetrations and sterility preservation. The valve of container should be made in aluminum, consists of single lid, protected by round cup,only in lid to assure perfect tightening during transport. One piece gasket. Inner SS basket with lid should be made up of AISI 304 grade stainless steel, perforated sheet with two handles. containers should have atleast 7 years of warranty. comply with EN ISO 17665-1, 16707-1 standards.	20	piece
1.106	Sterilization containers with compatible SS basket for steam sterilization : Dimesion of container in mm: 3/4 STU (460 X 300 X 260), inner basket measuring (400x240x180): stainless steel container should be AISI 304 grade, an austentic stainless steel with exceptional mechanical resistance and highly resistant corrosion. Container lid should have to be provided with a protected valve system that assures a mechanical control of steam penetrations and sterility preservation. The valve of container should be made in aluminum, consists of single lid, protected by round cup,only in lid to assure perfect tightening during transport. One piece gasket. Inner SS basket with lid should be made up of AISI 304 grade stainless steel, perforated sheet with two handles. containers should have atleast 7 years of warranty. comply with EN ISO 17665-1, 16707-1 standards.	20	piece
1.107	Sterilization containers with compatible SS basket for steam sterilization : Dimesion of container in mm: 1/2 STU (300 X 300 X 260), inner basket measuring (240x240x90): stainless steel container should be AISI 304 grade, an austentic stainless steel with exceptional mechanical resistance and highly resistant corrosion. Container lid should have to be provided with a protected valve system that assures a mechanical control of steam penetrations and sterility preservation. The valve of container should be made in aluminum, consists of single lid, protected by round cup,only in lid to assure perfect tightening during transport. One piece gasket. Inner SS basket with lid should be made up of AISI 304 grade stainless steel, perforated sheet with two handles. containers should have atleast 7 years of warranty. comply with EN ISO 17665-1, 16707-1 standards.	20	piece
1.108	Wired Mesh Tray: stainless steel wire mesh tray size 17" x 10" x 3" (l x b x h) made up of 16 Gauge wire of 1.5mm thickness. Borders and movable handle should be made up of 4mm thickness. Edges should be smooth. It should be designed for all phases of instruments processing (washing, disinfection, ultra sonic cleaning, sterilization, etc.) chemical resistant.	30	piece
1.109	Wired Mesh Tray: stainless steel wire mesh tray size 12" x 8" x 1.5" (l x b x h) made up of 16 Gauge wire of 1.5mm thickness. Borders and movable handle should be made up of 4mm thickness. Edges should be smooth. It should be designed for all phases of instruments processing (washing, disinfection, ultra sonic cleaning, sterilization, etc.) chemical resistant.	400	piece
1.110	SS Containers for surgical instruments and trays: High quality stainless steel (SS 304/316) containers measuring 100cm x 60cm x 30 cm (L X BX H) for keeping surgical instruments and trays in a enzymatic solution for cleaning. Catalogue to be provided.	10	piece

1.111	Tray Liner: For keeping in the bottom of the instruments tray to absorb the moisture during the steam sterilisation process. size 55cm x 18cm	30000	piece
1.112	Reusable Silicon Mesh Tray: For keeping in the bottom of the instruments tray, should be compatible with steam sterilisation process. size 12" x 8".	400	piece
1.113	Diaphragm Plate with gasket (cup type) for steam sterilizer	50	piece
1.114	I.V. Butterfly Needle with catheter: sterile, single use scalp vein (winged needle) infusion set. Needle Size: 0.7mm x 19mm (25G),. It should have flexible kink resistant extension tube (200mm +/-5%) with protective needle sheath. Shelf life : ≥ 3 years, Standard: IS 16097 Latest version.	5000	piece
2	NON-CONSUMABLES ITEMS		
2.115	Hand Dryer Should be wall mount type. Should have infrared sensor for automatic detection of hands. Should have brushed 304 SS finish. Motor should be at least 2300 W and 7500 RPM. Dryer should deliver the flow of 7300 LFM. Should work on 230V, 50 Hz power supply. Should supply with all accessories such as clamps for mounting (Brochure to be provided).	2	piece
2.116	Drying Cabinet for CSSD Should be automatic in operation. Inner chamber should be made up of stainless steel and outer chamber should be of epoxy painted CRCA sheets. Should have heaters of minimum 1.5 KW. There should be provision for setting the drying temperature and drying time. Approximate capacity 200L or more with shelves to keep trays (Brochure to be provided).	1	piece
2.117	Colour coding tag system for surgical Instruments with accessories (Brochure to be provided) Number of instruments need to be tagged: 6000	1	piece
2.118	Stainless Steel trolley for CSSD Specifications: Material S.S. 304 grade with heavy duty castor wheels. Size: 60" x 30" x 40"(LXBXH). 3 shelves and castor wheels of 6" diameter	2	piece
2.119	Stainless Steel Table for CSSD stainless steel table with 1000 kg weight bearing capacity measuring 7 ft. x 2 ft. x 1" (l x b x thickness) table top. Table height should be 2.5ft. 4 to 6 interconnected legs, each leg should be 3" x 2".	2	piece
2.120	Water Purifier 120 litres stainless steel, 2 faucets (normal & hot water), 7 stage RO water purifier system with capacity of 60 LPH.	2	piece
2.121	Ultrasonic Cleaner (Ultrasonic Bath) for surgical instrument: Digital microprocessor controlled, stainless steel (SS304/316, 16G & outer SS) ultrasonic surgical instrument cleaner 40 litres capacity, 40 ± 5 KHz operating frequency, impulses through wash water containing detergents and electrical heating. Complete Data / Brochure / picture should be submitted for evaluation.	2	piece
2.122	Document Scanner Sheetfed digital scanner, resolution upto 600 dpi, scan speed of upto 65 ppm / 130 ipm, 2 scan mode (simplex / duplex), scan file format including PDF, word, excel, power point, JPEG, PNG, etc., both grey scale mode and colour scanning. Complete Data / Brochure / picture should be submitted for evaluation.	2	piece
2.123	Sterile Storage Cabinets for CSSD Specifications: Material S.S. 304 grade. Size: 40"x 15"x 60"(LXBXH), 4 shelves, double doors with lockable facility and UV lights.	2	piece
2.124	Digital Weighing Machine: Stainless steel 600 x 600mm platform size weighing machine with LED display least count upto 10 gm to maximum 150 kg weighing capacity with one side protective side bar.	2	piece
2.125	ATP complete handheld device: A reliable, quick, easy use ATP meter measuring RLU to check for microbial contamination to measure the efficacy of cleaning protocols on surgical instruments and non-critical surfaces. Firm suppose to be provide compatible 50 swab free of cost along with the ATP meter	1	piece

Minimum 10 pieces of samples along with necessary documents mentioned in the specification for each item quoted should be submitted to the "Officer-in-Charge of CSSD, Central Sterile Supplies Department (CSSD), JIPMER, Dhanvantari Nagar, Puducherry - 605 006" within ten days of the closing of online submission of bids failing which the bids will be rejected.

Terms and Condition

1. JIPMER SUPPLY NOT FOR SALE should be printed / inscribed on all units as well as packages.
2. DOOR DELIVERY of the items, i.e., delivery to the department of CSSD is must.
3. Samples labeled each item with tender Serial no. must be sent to the “Officer-in-Charge of Central Sterile Supplies Department (CSSD), JIPMER, Puducherry - 605 006”, within **ten days of the closing of online submission of bids** failing which the bids will be rejected.
4. Supplied items should have a minimum shelf life of two years.
5. Samples of non-consumable costly items above Rs.1500/- if not selected should be collected by the concerned firm before the date of opening of next tender. No claim shall be entertained after this period. For the Non-Consumables Items, complete brochure, Data Sheet, Picture of the items and all necessary certificates to be sent to the above mentioned address.
6. Items having ISI / ISO similar certification or equivalent are likely to get preference.
7. Selected firms should supply the items in conformance with samples and specifications within the stipulated period with clear details in the packing such as quantity, date of manufacture, expiry date and batch nos if any, PACK SIZE should be maintained & delivery quantity should be as per CSSD delivery schedule.
8. Furniture samples of each item 03 Nos should be submitted. Samples labeled each item with tender Serial No. must be sent to the Officer-in-Charge of CSSD, JIPMER, Puducherry, before the last date, if failing which the tender will not be consider as lowest.
9. Expiry of goods if any during the supply period, the expiry goods should be replaced with long period with free of cost to CSSD, JIPMER, Puducherry.

Section-III

GENERAL INSTRUCTIONS TO TENDERERS

1. **Period of contract:**

The contract shall initially be for a period of **ONE year w.e.f date of opening of price bid that may be extended to 6 months**. The rates approved shall remain unchanged during the period of contract.

2. **Earnest Money Deposit (EMD)**

- i. The tenderer shall furnish along with its tender, Earnest Money for an amount of **Rs.5000/- (Rupees five thousand only)**. The Earnest Money is required to protect the purchaser against the risk of the tenderer's unwarranted conduct.
- ii. The tenderers who are currently registered and, also, will continue to remain registered during the tender validity period as Micro and Small Enterprises (MSEs) (**Only Manufacturers** for the items being quoted) as defined in MSE Procurement Policy issued by Department of Micro, Small and Medium Enterprises (MSME) or with National Small Industries Corporation, shall be eligible for exemption from EMD. In case the tenderer falls in this category, it should furnish copy of its valid registration details (with MSME or NSIC, as the case maybe). **Traders and service providers are not exempted from EMD.**
- iii. The registered vendors of JIPMER are exempted for EMD as per JIPMER purchase guidelines. In case the tenderer falls in this category, it should furnish copy of its valid registration details.
- iv. The earnest money shall be denominated in Indian Rupees only and paid through SBI collect only.
- v. The earnest money shall be valid for a period of forty-five (45) days beyond the validity period of the tender.
- vi. Unsuccessful tenderers' earnest money will be returned to them without any interest after expiry of the tender validity period, but not later than thirty days after conclusion of the resultant contract. Successful bidder's earnest money will be returned without any interest, after receipt of performance security from that tenderer.
- vii. Earnest Money is required to protect the purchaser against the risk of the Tenderer's conduct which would warrant the forfeiture of the EMD. Earnest money of a tenderer will be forfeited if the tenderer withdraws or amends its tender or impairs or derogates from the tender in any respect within the period of validity of its tender or if it comes to notice that the information/documents furnished in its tender is incorrect, false, misleading or forged without prejudice to other rights of the purchaser. The successful bidder's earnest money will be forfeited without prejudice to other rights of Purchaser if it fails to furnish the required performance security within the specified period.

3. **Eligibility conditions of bidders:**

- i. For items where the estimated value of annual procurement exceeds Rs 5 lakhs, the conditions as stated in Public Procurement (Preference to Make in India), order 2017 of MoC and I (DIPP), Govt. Of India, as amended from time to time, shall apply. For such items, the Tenderer must be a domestic manufacturer that is a 'Class-I local supplier' or a 'Class-II local supplier' as defined under the above mentioned orders. For determining the eligibility of bidders to participate in the tender for such items where the estimated value of annual procurement of the item is more than Rs.5 lakhs bidders must furnish in tabular form as given in Appendix A the percentage of local content added and the place in this country where it was added, for each of such items, failing which the quote for such items where the required information is not provided in the Appendix as mentioned above, will be rejected summarily as being nonresponsive. In case the manufacturer does not quote directly, they may authorize an agent as per proforma of Manufacturer authorization form as given in the Tender enquiry document to quote and enter into a contractual obligation.
- ii. In compliance with order (Public Procurement No.1) No. 6/18/2019-PPD dated 23rd July 2020 issued by the Public Procurement Division, Dept. of Expenditure, Min of Finance under Rule 144(xi) of GFR 2017 any bidder 'from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority.
 - a. "Bidder" (including the term 'tenderer', 'consultant' or 'service provider' in certain contexts) means any person or firm or company, including any member of a consortium or joint venture (that is an association of several persons, or firms or companies), every artificial juridical person not falling in any of the descriptions of bidders stated hereinbefore, including any agency branch or office controlled by such person, participating in a procurement process.
 - b. "Bidder from a country which shares a land border with India" for the purpose of this Order means: -
 - I. An entity incorporated, established or registered in such a country; or
 - II. A subsidiary of an entity incorporated, established or registered in such a country; or
 - III. An entity substantially controlled through entities incorporated, established or registered in such a country; or
 - IV. An entity whose beneficial owner is situated in such a country; or
 - V. An Indian (or other) agent of such an entity; or
 - VI. A natural person who is a citizen of such a country; or
 - VII. A consortium or joint venture where any member of the consortium or joint venture falls under any of the above
 - c. The beneficial owner for the purpose of above will be as under:
 - I. In case of a company or Limited Liability Partnership, the beneficial owner is the natural person(s), who, whether acting alone or together, or through one or more juridical person, has a controlling ownership interest or who exercises

control through other means.

Explanation—

- “Controlling ownership interest” means ownership of or entitlement to more than twenty-five per cent, of shares or capital or profits of the company;
 - "Control" shall include the right to appoint majority of the directors or to control the management or policy decisions including by virtue of their shareholding or management rights or shareholders agreements or voting agreements;
 - II. In case of a partnership firm, the beneficial owner is the natural person(s) who, whether acting alone or together, or through one or more juridical person, has ownership of entitlement to more than fifteen percent of capital or profits of the partnership.
 - III. In case of an unincorporated association or body of individuals, the beneficial owner is the natural person(s), who, whether acting alone or together, or through one or more juridical person, has ownership of or entitlement to more than fifteen percent of the property or capital or profits of such association or body of individuals.
 - IV. Where no natural person is identified under (a) or (b) or (c) above, the beneficial owner is the relevant natural person who holds the position of senior managing official;
 - V. In case of a trust, the identification of the beneficial owner(s) shall include the identification of the author of the trust, the trustee, the beneficiaries with 15% or more interest in the trust and any other natural person exercising ultimate effective control over the trust through a chain of control or ownership.
- d. An Agent is a person employed to do any act for another, or to represent another in dealings with third person.
- e. In case of turnkey contracts, the successful bidder shall not be allowed to sub-contract works to any contractor from a country which shares a land border with India unless such contractor is registered with the Competent Authority.
- f. Bidders must certify compliance with the above mentioned provisions in the tender form as per **Section VII**.
- g. The bidder should have successfully completed the delivery of the item of at least 25% of the quantity required in this tender to a government hospital in the last 3 years.
- h. The bidder or the manufacturer must have an average annual turnover of at least Rs.1,00,00,000/- (Rupees one crore only) each during the last three financial years i.e. 2019-20, 2020-21, 2021-21.
- i. Bidder should have ISO Certification.

4. Purchase Preference

- i. The Procurement of goods and services under this e-tender will be regulated as per the applicable provisions of Public Procurement (Preference to Make in India), order 2017 of MoC and I (DIPP), Govt. Of India, as further amended by orders of even number dated 28.05.2018, 29.05.2019, 04.06.2020 and 16.09.2020. Salient portions of the

order are reproduced in Appendix A of this tender document by way of information. Bidders are advised to see the original orders and satisfy themselves that they qualify to participate in the tender. Bidders who are claiming eligibility to participate in this tender must submit a certificate in format given in Appendix A along with documentary evidence in support of their claim wherever necessary failing which their bid will be summarily rejected. The purchaser reserves the right to give preference to the 'Class-I local supplier'.

- ii. The Purchaser reserves the right to give the purchase preference to small-scale sectors, Micro and small scale enterprises etc. as per the instruction in vogue while evaluating, comparing and ranking the responsive Tenders as given in the MSMED Act 2006 reproduced below:
 - a. In exercise of powers conferred in Section 11 of the Micro, Small and Medium Enterprises Development (MSMED) Act 2006, the Government has notified a new Public Procurement Policy for Micro & Small Enterprises effective from 1st April 2011. The policy mandates that 25% of procurement of annual requirement of goods and services by all Central Ministries/Public Sector Undertakings will be from the micro and small enterprises. The Government has also earmarked a sub-target of 4% procurement of goods & services from MSEs owned by SC/ST entrepreneurs out of above said 25% quantity.
 - b. In accordance with the above said notification, the participating Micro and Small Enterprises (MSEs) in a tender, quoting price within the band of L1+15% would also be allowed to supply a portion of the requirement by bringing down their price to the L1 price, in a situation where L1 price is from someone other than an MSE. Such MSEs would be allowed to supply up to 25% of the total tendered value. In case there are more than one such eligible MSE, the 25% supply will be shared equally. Out of 25% of the quantity earmarked for supply from MSEs, 5% quantity is earmarked for procurement from MSEs owned by SC/ST entrepreneurs. However, in the event of failure of such MSEs to participate in the Tender process or meet the tender requirements and the L1 price, the 5% quantity earmarked for MSEs owned by SC/ST entrepreneurs will be met from other participating MSEs.
 - c. The MSEs fulfilling the prescribed eligibility criteria and participating in the Tender shall enclose with their tender a copy of their valid registration certificate with District Industries Centres or Khadi and Village Industries Commission or Khadi and Village Industries Board or Coir Board or National Small Industries Corporation or any other body specified by Ministry of Micro and Small enterprises in support of their being an MSE, failing which their tender will be liable to be ignored.
 - d. Special provision for Micro and Small Enterprise owned by women: – Out of the total annual procurement from Micro and Small Enterprises, 3 per cent from within the 25 per cent target shall be earmarked for procurement from Micro and Small Enterprises owned by women.
 - e. Note: “If the bidder is a MSME, it shall declare in the bid document the Udyog Aadhar Memorandum Number issued to it under the MSMED Act, 2006. If a MSME bidder do not furnish the UAM Number along with bid documents, such MSME unit will not be eligible for the benefits available under Public Procurement Policy for MSEs Order 2011.

5. **Demonstration of sample**

The tenderer must give a sample within **ten** days of the closing of online submission of bids failing which the bids will be rejected.

6. **Instructions for the filling the tender form**

- i. E-Tender form shall be completed in all respect, signed in full and stamped at appropriate places and initialed and stamped on all remaining pages. Incomplete or e-tenders without tender processing fee, EMD, Make-in-India self-certification as in **Appendix-A** only by original manufacturer, **GFR 144 (xi) compliance certificate** as in **Section-IX** shall be treated as invalid.
- ii. Bidders have to ensure that all the documents are properly filled.
- iii. Conditional tenders are liable to be rejected.
- iv. Bids received and found valid will be evaluated by JIPMER to ascertain the complete work/services under the specification and documents. The bidder should take care to submit all the information sought by JIPMER in prescribed formats.
- v. Incomplete bids, bids in paper format, conditional bids, telephonic bids or tenders submitted after the due date and time will not be considered and summarily rejected. Vendors are, therefore, advised to submit their bids well on time.
- vi. **The bidder can quote for one or more items mentioned in the list. Bidder has to give all details (HSN, MSME, Make-in-India, make/brand, model, pack size and remark) mentioned in BOQ for all quoted items, failure of that the bid will be rejected summarily.**

7. **Submission of tenders**

The bidders must ensure that they submit the **on-line bids** within the scheduled closing date & time.

8. **Late Tender:**

There is NO PROVISION of uploading late tender beyond stipulated date & time in the e- tendering system.

9. **Alteration and Withdrawal of Tender**

- i. The bidder, after submitting its bid, is permitted to alter/modify its bid, within the deadline for submission of bids. Alterations/modifications to bids received after the prescribed deadline will not be possible on the e-tender portal.
- ii. No tender should be withdrawn or modified after the deadline for submission of tender and before expiry of the tender validity period. If a bidder withdraws or modifies the tender during this period, it will result in forfeiture of the EMD furnished by the bidder in its bid.

10. Preparation of e-tenders

This is a Two-Bid Tender system, consisting of the **Techno-Commercial Bid and Price Bid** that are to be uploaded in the prescribed formats in the e-tendering portal. The tender(s) shall only be submitted online as mentioned below:

I. **Techno-commercial Bid shall comprise**

- a. Fee Cover
 - i. **E-tender Processing fee of Rs.590/-** (Rupees five hundred and ninety only) inclusive of 18% GST payment **receipt** duly self-attested and rubber stamped should be uploaded.
 - ii. Scanned copy in pdf format of **EMD receipt** or, if EMD exemption is claimed, **copy of valid registration** details proving that the bidder is a Micro or Small enterprise (**Only Manufacturer** for the items being quoted) or is registered as a Small-Scale Industry with MSE or Valid registration certificate if registered with JIPMER, as the case may be, should be uploaded. **Traders and service providers are not exempted from EMD.**
- b. In the cover named “**Prequal**” the scanned copy in pdf format of the following documents are to be uploaded:
 - i. The **Bidder’s Profile** as per in **Section-V** must be downloaded duly filled signed and stamped and uploaded.
 - ii. **Tender Form** as in **Section-VII** must be downloaded duly filled signed and stamped and uploaded.
 - iii. **Copies of abridged Annual report of last 03 years** (Income tax return acknowledgement, Assets and Liabilities, Balance sheet and Profit & Loss Account) must be uploaded as a single PDF file
 - iv. Copy of Self Certified **GST registration certificate** and Copy of **PAN Card** must be uploaded as a single PDF file.
 - v. **Bank Details** (Beneficiary name, Bank name, Account number, IFSC code, Branch address on letterhead).
 - vi. **Check list** as in **Section-VIII** in the prescribed format duly filled and signed must be uploaded.
- c. In the cover named “**Technical**” the scanned copy in pdf format of the following documents are to be uploaded:
 - i. **Copies of Supply orders/Completion certificate** in support of Eligibility condition 1 and 1.
 - ii. **Manufacturer’s Authorization letter** in company letterhead in format in given **Section-VI**
 - iii. Competent Authority under **GFR 144 (xi) (mandatory)** a copy of the same or **GFR 144(xi) compliance certificate** as in **Section-IX** must be uploaded.

- iv. Self-certification for supporting the claim to be a local supplier under the “Public Procurement preference to Make in India” order in format as in **Appendix-A (mandatory)** mentioning percentage of value addition in India and address of value addition for each item quoted, by Original Manufacturer and any other document that the bidder wishes to submit as a single PDF file.
- v. A file mentioning the list of items with their Make/Brand, Model and Pack size **WITHOUT PRICE** for which bidder is quoting must be uploaded as a single PDF file.

II. Price Bid:

Prices are to be quoted in the prescribed Price Bid format provided in the e-tender portal using the BOQ template only. The price should be quoted for the **accounting unit** indicated in the e-tender document.

Note:

- i) The bidder has to be diligent while filling up the Techno-Commercial Bid and Price Bid provided in prescribed formats and must not tamper with the contents of the sheets.
- ii) Bidders must ensure that the documents uploaded in pdf format are legible.
- iii) It is the responsibility of bidder to go through the Tender document to ensure furnishing all required documents in addition to above, if any.
- iv) The Make in India self-certification as in Appendix-A must be given and uploaded by original manufacture. The distributor can upload the Appendix-A given by original manufacture.
- v) ITE- Item-wise Eligibility Sheet should be downloaded, the items that the bidder wishes to quote must be selected as “Eligible”, and this “ITE file” must also be uploaded for the price bid to be considered by the system. The selected items will be displayed once uploaded and the bidder can verify that all items he wishes to quote for, are present in the list
- vi) A person signing (manually or digitally) the tender form or any documents forming part of the contract on behalf of another shall be deemed to warrant that he has authority to bind such other persons and if, on enquiry, it appears that the persons so signing had no authority to do so, the purchaser may, without prejudice to other civil and criminal remedies, cancel the contract and hold the signatory liable for all cost and damages.
- vii) A bid, which does not fulfill any of the above requirements and/ or give evasive information/reply against any such requirement, shall be liable to be ignored.
- viii) Tender sent by fax/telex/cable shall be ignored.

11. Digital Signing of Tender

The tenderers shall submit their tenders as per the instructions contained as above. Tenders shall be uploaded with all relevant tender documents in the prescribed format. The relevant tender documents should be uploaded by an authorized person having Class 3 digital signature certificate.

12. Tender currencies.

The tender shall be quoted only in INR.

13. Additional information and instruction on GST:

If the Tenderer desires to ask for GST or any other taxes to be paid extra, the same must be specifically stated. In the absence of any such stipulation, the price will be taken inclusive of such taxes and no claim for the same will be entertained later. The rate of GST quoted in the tender shall be taken for price comparison. However, the rate of GST quoted in the tender or the actual rate of GST applicable, whichever is lower shall be payable by the purchaser. The supplier can charge a higher GST than quoted in the tender only if the rate of GST was revised by the government after the tender closing date.

14. Tender opening

- i. The Tender Inviting Authority will open the e-tenders at the specified date and time and at the specified place as indicated in the NIT. In case the specified date of tender opening falls on / is subsequently declared a holiday or closed day for the Tender Inviting Authority, the tenders will be opened at the appointed time and place on the next working day.
- ii. Authorized representatives of the tenderers, who have submitted tenders on time, may attend the tender opening provided they bring with them letters of authority from the corresponding tenderers. The tender opening official(s) will prepare a list of the representatives attending the tender opening. The list will contain the representatives' names & signatures and corresponding tenderers' names and addresses.
- iii. This being a Two-Bid Tender system, the **Techno-Commercial Tenders** are to be opened in the first instance, at the prescribed time and date as indicated in NIT. These Tenders shall be scrutinized and evaluated by the competent committee/ authority with reference to parameters prescribed in the tender document. During the Techno-Commercial Tender opening, the tender opening official(s) will read the salient features of the tenders like brief description of the goods offered, delivery period, EMD and any other special features of the tenders, as deemed fit by the tender opening official(s). Thereafter, in the second stage, the Price bids of only the Techno - Commercially acceptable offers (as decided in the first stage) shall be opened for further scrutiny and evaluation on a date notified after the evaluation of the Techno-Commercial tender.

15. Scrutiny and evaluation of tenders

A. Basic Principle

Tenders will be evaluated on the basis of the terms & conditions already incorporated in the tender enquiry document, based on which tenders have been received and the terms, conditions etc. mentioned by the tenderers in their tenders. No new condition will be brought in while scrutinizing and evaluating the tenders.

B. Scrutiny of Tenders

- i. The Purchaser will examine the Tenders to determine whether they are complete, whether any computational errors have been made, whether required sureties have

- been furnished and, whether the documents uploaded are in legible form.
- ii. The Purchaser's determination of a Tender's responsiveness is to be based on the contents of the tender itself without recourse to extrinsic evidence.
 - iii. The tenders will be scrutinized to determine whether they are complete and meet the essential and important requirements, conditions etc. as prescribed in the tender document. The tenders, which do not meet the basic requirements, are liable to be treated as non-responsive and will be summarily ignored.
 - iv. The following are some of the important aspects, for which a tender shall be declared non-responsive during the evaluation and will be ignored;
 - i. Tender validity is shorter than the required period.
 - ii. Non-submission of EMD receipt or EMD exemption certificate.
 - iii. Non-submission of receipt of tender processing fee.
 - iv. Non-submission of self-certification in format as given in Appendix-A only by original manufacturer, for determining eligibility to participate in the tender under the "Public Procurement preference to Make in India" order.
 - v. Non-submission of GFR-144 (xi) compliance certificate.
 - vi. Tenderer has not agreed to give the required performance security of required amount in an acceptable form.
 - vii. Non-submission of samples within ten days of the closing of online submission of bids
 - viii. Poor/ unsatisfactory past performance.
 - ix. Tenderers who stand de-registered/banned/blacklisted by any Central Govt. Ministries/Departments/Hospitals/Institutes.
 - x. Tenderer is not eligible as per tender conditions.
 - xi. Tenderer has not quoted for the entire quantity as specified in the List of Requirements/ BOQ for the item quoted.
 - xii. Non-submission of all details of quoted items (HSN, MSME, Make-in-India, make/brand, model, pack size and remark).
 - xiii. Tenderer has not agreed to other essential condition(s) specially incorporated in the tender enquiry, like delivery terms, delivery schedule, terms of payment, liquidated damages clause, warranty clause, dispute resolution mechanism applicable law.

16. Minor Informality/Irregularity/Non-Conformity

If during the preliminary examination, the purchaser find any minor informality and/or irregularity and/or non-conformity in a tender, the purchaser may waive the same provided it does not constitute any material deviation and financial impact and, also, does not prejudice or affect the ranking order of the tenders. Wherever necessary, the purchaser will convey its observation on such 'minor' issues to the tenderer by registered/speed post etc. asking the tenderer to respond by a specified date. If the tenderer does not reply by the specified date or gives evasive reply without clarifying the point at issue in clear terms, that tender will be liable to be ignored.

17. Award of work

- i. The selection of the agency will be at the sole discretion of the JIPMER who reserves its right to accept or reject any or all the proposals without assigning any reason thereof.
- ii. The lowest bid will be evaluated on the basis of annual estimated amount as quoted by bidders in the quotation including taxes.
- iii. Upon evaluation of offers the decision on the award of contract will be intimated to the successful bidder.
- iv. The annual estimate is given only as an indication. The actual quantity procured may increase or decrease. No assurance is given that the quantity stated will actually be procured.

18. Code of Integrity in Public Procurement; Misdemeanours and Penalties

18.1 Code of Integrity

Procuring authorities as well as bidders, suppliers, contractors, and consultants - should observe the highest standard of ethics and should not indulge in following prohibited practices, either directly or indirectly, at any stage during the Tender Process or during the execution of resultant contracts:

1. **“Corrupt practice”** - making offer, solicitation or acceptance of a bribe, reward or gift or any material benefit, in exchange for an unfair advantage in the Tender Process or to otherwise influence the Tender Process;
2. **“Fraudulent practice”** - any omission or misrepresentation that may mislead or attempt to mislead so that financial or other benefits may be obtained or an obligation avoided. Such practices include a false declaration or false information for participation in a tender process or to secure a contract or in the execution of the contract;
3. **“Anti-competitive practice”** - any collusion, bid-rigging or anti-competitive arrangement, or any other practice coming under the purview of the Competition Act, 2002, between two or more bidders, with or without the knowledge of the Procuring Entity, that may impair the transparency, fairness, and the progress of the Tender Process or to establish bid prices at artificial, non-competitive levels;
4. **“Coercive practice”** - harming or threatening to harm persons or their property to influence their participation in the Tender Process or affect the execution of a contract;
5. **“Conflict of interest”** –participation by a bidding firm or any of its affiliates who are either involved in the Consultancy Contract to which this procurement is linked; or if they are part of more than one bid in the procurement; or if their personnel have a relationship or financial or business transactions with any official of procuring entity who are directly or indirectly related to tender or execution process of contract; or improper use of information obtained by the (prospective) bidder from the Procuring Entity with an intent to gain unfair advantage in the Tender Process or for personal gain;

6. **“Obstructive practice”** - materially impede procuring entity’s investigation into allegations of one or more of the above mentioned prohibited practices either by deliberately destroying, falsifying, altering; or by concealing of evidence material to the investigation; or by making false statements to investigators and/ or by coercive practices mentioned above, to prevent it from disclosing its knowledge of matters relevant to the investigation or from pursuing the investigation; or by impeding the Procuring Entity’s rights of audit or access to information;

18.2 Obligations for Proactive Disclosures:

Procuring authorities, bidders, suppliers, contractors, and consultants are obliged under this Code of Integrity to suo-moto proactively declare any conflict of interest (coming under the definition mentioned above - pre-existing or as and as soon as these arise at any stage) in any Tender Process or execution of the contract. Failure to do so shall amount to a violation of this code of integrity. Any bidder must declare, whether asked or not in a bid-document, any previous transgressions of such code of integrity during the last three years or of being under any category of debarment by the Central Government or by the Ministry/ Department of the Procuring Organisation from participation in Tender Processes. Failure to do so shall amount to a violation of this code of integrity.

18.3 Misdemeanours and Penalties

The following shall be considered misdemeanours - if a bidder/ contractor either directly or indirectly, at any stage during the Tender Process or during the execution of resultant contracts:

- 1) commits any of the following misdemeanours:
 - a) violates the code of Integrity mentioned in GCC-clause 13.1 or GCC-Clause 10.1.6 (Fall clause) or the Integrity Pact if included in the Tender/ Contract;
 - b) any other misdemeanour, e.g., supply of sub-standard quality of material/ services/ work or non-performance or abandonment of contract or failure to abide by ‘Bid Securing Declaration’.
- 2) commits any of the following misdemeanours:
 - a) has been convicted of an offence:
 - (i) under the Prevention of Corruption Act, 1988; or
 - (ii) the Indian Penal Code or any other law for the time being in force for causing any loss of life or property or causing a threat to public health as part of the execution of a public procurement contract.
 - b) is determined by the Government of India to have doubtful loyalty to the country or national security consideration.
 - c) Employs a government servant, who has been dismissed or removed on account of corruption or employs a non-official convicted for an offence involving corruption or abetment of such an offence, in a position where he could corrupt government servants or employs a government officer within one year of his retirement, who has had business dealings with him in an official capacity before retirement.

18.4 Penalties for Misdemeanours

Without prejudice to and in addition to the rights of the Procuring Entity to other remedies as per the Tender-documents or the contract, If the Procuring Entity concludes that a (prospective) bidder/ contractor directly or through an agent has committed a misdemeanour in competing for the tender or in executing a contract, the Procuring Entity shall be entitled, and it shall be lawful on his part to take appropriate measures, including the following:

18.4.1 if his bids are under consideration in any procurement

- 1) Enforcement of Bid Securing Declaration in lieu of forfeiture or encashment of Bid Security.
- 2) calling off of any pre-contract negotiations, and;
- 3) rejection and exclusion of Bidder from the Tender Process

18.4.2 if a contract has already been awarded

- 1) Termination of Contract for Default and availing all remedies prescribed thereunder;
- 2) Encashment and/ or Forfeiture of any contractual security or bond relating to the procurement;
- 3) Recovery of payments including advance payments, if any, made by the Procuring Entity along with interest thereon at the prevailing rate (MIBID - Mumbai Interbank Bid Rate)

18.4.3 Remedies in addition to the above:

In addition to the above penalties, the Procuring Entity shall be entitled, and it shall be lawful on his part to:

- 1) File information against Bidder or any of its successors, with the Competition Commission of India for further processing, in case of anti-competitive practices;
- 2) Initiate proceedings in a court of law against Bidder or any of its successors, under the Prevention of Corruption Act, 1988 or the Indian Penal Code or any other law for transgression not addressable by other remedies listed in this subclause.
- 3) Remove Bidder or any of its successors from the list of registered suppliers for a period not exceeding two years. Suppliers removed from the list of registered vendors or their related entities may be allowed to apply afresh for registration after the expiry of the period of removal.
- 4) Initiation of suitable disciplinary or criminal proceedings against any individual or staff found responsible.
- 5) Debar, a bidder/ contractor from participation in future procurements without prejudice to Procuring Entity's legal rights and remedies. Debarment shall automatically extend to all the allied firms of the debarred firm. In the case of Joint Venture/ consortium, all its members shall also stand similarly debarred:
 - a) A Ministry/ Department (or any of its CPSUs, attached offices, autonomous bodies) may debar a bidder or any of its successors from participating in any Tender Process undertaken by all its procuring entities for a period not exceeding two years commencing from the date of debarment for misdemeanours listed in sub-clause GCC 13.3 -1) above. The Ministry/Department shall maintain such a list which shall also be displayed on their website.

- b) Central Government (Department of Expenditure (DoE), Ministry of Finance) may debar a bidder or any of its successors from participating in any Tender Process undertaken by all its procuring entities for a period not exceeding three years commencing from the date of debarment for misdemeanours listed in sub-clause GCC 13.3 - 2) above. DoE shall maintain such a list which shall be displayed on Central Public Procurement Portal (CPPP).

Section-IV
GENERAL CONDITIONS OF CONTRACT

1. Price of goods

The rate quoted in the e-tender will be fixed for the whole contract period.

2. Technical Specifications and Standards

The Goods & Services to be provided by the supplier under this contract shall conform to the technical specifications mentioned in “Technical Specification” under **Section II**.

3. Terms of Delivery

- i. Goods of required quantity said by Officer-in-Charge, CSSD should be initially delivered by the supplier within 30 days of issue of supply order whereas the bulk quantities should be supplied as per the time schedule to be given by CSSD. Please note that the time shall be the essence of the contract.
- ii. Any unexcused delay by the supplier in maintaining its contractual obligations towards delivery of goods and performance of services shall render the supplier liable to any or all of the following sanctions:
 - a) Imposition of liquidated damages,
 - b) Forfeiture of its performance security and
 - c) Termination of the contract for default.

4. Liquidated Damages

If the supplier fails to deliver any or all of the goods or fails to perform the services within the time frame(s) incorporated in the contract, the Purchaser/Consignee shall, without prejudice to other rights and remedies available to the Purchaser/Consignee under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods, installation, commissioning and/or services until actual delivery or performance subject to a maximum of 10% of the contract price. Once the maximum is reached Purchaser/Consignee may consider termination of the contract as per GCC 8. Since the Liquidated damages are in virtue of non-performance of services, it will attract GST or any other applicable taxes which in turn shall be deducted from the bidder.

5. Performance Security

The successful bidder shall have to deposit an amount 3% of the value of contract as Performance Security Deposit (PSD) within two weeks after award of contract, through SBI Collect available on JIPMER website. In the event of any failure /default of the supplier with or without any quantifiable loss to the purchaser, the amount of the performance security is liable to be forfeited.

Subject to the condition mentioned above the Performance Security will be released without any interest to the supplier on completion of the supplier's all contractual obligations including the warranty obligations and extension of time (with or without Liquidated Damages).

6. Payment Procedure

The contractor shall submit bill in triplicate upon satisfactory supply of goods. Bill must be raised based on the rate quoted in e-tender. Every effort shall be made to ensure that the payment will be made within 45 days of submission of bill. No advance payment will be considered. TDS will be deducted as per provision of Income Tax Act, GST Acts and other statutes as relevant.

7. Risk Clause

The contractor shall at all times have standby arrangements for carrying out the work under the contract, in case of any failure of the existing arrangements. JIPMER reserves the right for termination of the contract at any time by giving 30 days written notice, if the items delivered are found to be unsatisfactory and also has the right to award the contract to the next higher bidder willing to supply the item at the cost, risk and responsibilities of contractor and excess expenditure incurred on account of this will be recovered by JIPMER from the contractor's Performance Security Deposit or pending bills or by raising a separate claim.

8. Termination clause:

During the period of agreement if it is found that the agency is not providing proper services, the JIPMER reserves rights to make the vendor forfeit the security deposit deposited with JIPMER or part thereof in favour of JIPMER and agreement will be terminated after giving 30 days' notice. Furthermore, in such situations, tender can be allotted to second lowest bidder and the difference in cost shall be recovered from the earlier vendor who is breach of the contract. In addition in case it is found that the supplier is charging by fraudulent means or indulging in criminal activities the contract will be terminated immediately.

9. Jurisdiction of the courts

Jurisdiction of the courts for settlement of disputes:- Jurisdiction for the settlements of disputes if any is Puducherry only.

In above mentioned conditions Director JIPMER reserves all the rights.

- 10. Storage facility:** For the essential medical items bulk in nature like gloves, syringes, IV sets, Gowns, etc., the selected bidder should establish a storage facility within 15 kms perimetry from JIPMER, Puducherry to keep these items of two months requirement for ready to supply anytime. Authorized person from JIPMER can visit the storage facility anytime to ensure uninterrupted supply of essential medical items in view of manufacturing delay, transportation delay, weather condition and natural calamities etc.

Section-V

BIDDER'S PROFILE

This form duly filled and signed by authorized representative of the bidder and the scanned copy must be uploaded online

1.	Name & Designation of the contact person	
1.	Name and Address of the Tenderer	
3.	Phone No a) Land line number (functional between 9 am and 5 pm)	
4.	Mobile No of contact person (available from 9 am to 6 pm)	
5.	Email ID of the Tenderer	
6.	Email ID of the contact person	
7.	Local supplier/Distributor in Chennai/ Puducherry or any other place (complete address must be written)	
8.	Manufacture Name	
9.	Manufacture Address	
10	Whether Tenderer is registered MSE (Manufacture the product quoted) (If registered MSE, submit copy of the Udyog Aadhaar certificate or Equivalent Certificate)	Yes / No
<p>If there is any change in the above details, I will immediately intimate you by speed post or fax or email</p> <p>I.....hereby declare that the details given above are true to the best of my knowledge and I have thoroughly read and understood the terms and conditions of the tender and shall abide by the rules,</p> <p style="text-align: right;">Signature (Name and Designation & Seal)</p> <p>Dated:</p>		

NB: This form must be duly filled in by an authorized person

Section-VI
MANUFACTURER'S AUTHORISATION FORM
(Letterhead)

Date: _____

To
The Director,
JIPMER, Puducherry

Dear Sir,

Ref: Your TE document No: _____ dated: _____

We, _____ who are proven and reputable manufacturers of _____ (name and description of the goods offered in the Tender) having factories at _____, hereby authorize Messrs. _____ (name and address of the agent) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also state that we are not participating directly in this tender for the following reason(s):

(Please provide reason here).

We further confirm that no supplier or firm or individual other than Messrs _____ (name and address of the above agent) is authorised to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us. We also hereby extend our full warranty, CMC/AMC as applicable as per the General Conditions of Contract, read with modification, if any, in the Special Conditions of Contract for the goods and services offered for supply by the above firm against this TE document. We also hereby confirm that we would be responsible for the satisfactory execution of contract placed on the authorised agent.

We also confirm that the price quoted by our agent shall not exceed the price which we would have quoted directly”

Yours faithfully,
[Signature with date, name, designation and Email]
for and on behalf of Messrs _____
[Name & address of the manufacturers]

Note:

- (1) This letter of authorization should be on the letter head of the manufacturing firm and should be signed by a person competent and having the power of attorney to legally bind the manufacturer.
- (2) Original letter may be sent.
- (3) The purchaser reserves the right to verify this document with its signatory.

SECTION-VII
TENDER FORM
(On Firms' Letter Head)

To
The Director
JIPMER, Puducherry 605006

Date: _____

Ref. Your TE document No. _____ dated _____

We, the undersigned have examined the above mentioned TE document, including amendment/corrigendum No. _____, dated _____ (if any), the receipt of which is hereby confirmed. We now offer to supply and deliver _____ (Description of goods and services) in conformity with your above referred document **for the sum as shown in the price schedules attached herewith and made part of this tender**. If our tender is accepted, we undertake to supply the goods and perform the services as mentioned above, in accordance with the delivery schedule specified in the List of Requirements.

We further confirm that, if our tender is accepted, we shall provide you with a performance security of required amount in an acceptable form in terms of GCC, Special Conditions of Contract", for due performance of the contract.

We have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India and we certify that this bidder is not from such a country/ from such a country and has been registered with the Competent Authority and a copy of the valid registration by the Competent Authority is attached as evidence of the same (Strike out what is not applicable). In case there are Turnkey works to be carried out this bidder will not sub-contract any work to a contractor from such countries unless such contractor is registered with the Competent Authority. We hereby certify that this bidder fulfills all requirements in this regard and is eligible to be considered.

We agree to keep our tender valid for acceptance as required in the GIT, Special Instructions to Tenderers" or for subsequently extended period, if any, agreed to by us. We also accordingly confirm to abide by this tender up to the aforesaid period and this tender may be accepted any time before the expiry of the aforesaid period. We further confirm that, until a formal contract is executed, this tender read with your written acceptance thereof within the aforesaid period shall constitute a binding contract between us.

We further understand that you are not bound to accept the lowest or any tender you may receive against your above-referred tender enquiry.

We confirm that we do not stand deregistered/banned/blacklisted by any Govt. Authorities in the last 7 years.

We confirm that we fully agree to the terms and conditions specified in above mentioned TE document, including amendment/ corrigendum if any.

(Signature with date)
(Name and designation)

Duly authorized to sign tender for and on behalf of

Section-VIII

CHECK LIST FOR SUBMISSION OF TENDER

(To be filled by the tenderer and submitted along with the bid)

Sl. No.	Particular	Yes / No
1	i. Tender Processing Fee (Mandatory for all bidders) ii. EMD receipt or if EMD exemption, Valid MSE (Only Manufacturer) certificate or JIPMER Registration Certificate	
2	Copies of Supply orders/Completion certificate in support of Eligibility condition 2(ii).	
3	Copy of PAN Card	
4	Copy of ISO Certification of bidder, ISI/CE Certification of each product wherever asked	
5	Copies of last three years Income Tax Return statement with Balance sheets & Profit & Loss A/c i.e. 2019-20, 2020-21, 2021-21.	
6	Copy of GST Registration Certificate	
7	Authorization letter for signing tender documents if a person other than the Owner, Partner, Managing Director is signing/uploading the tender	
8	Tender form as in Section – VII duly signed and stamped	
9	Manufacturer's authorization form in format as in Section VI and	
10	Self-certification for supporting the claim to be a local supplier under the "Public Procurement preference to Make in India" order in format as in Appendix A	
11	A copy bidder's empanelment by the Competent Authority under GFR 144 (xi) or GFR 144 (xi) compliance certificate	
12	A PDF file containing list of all items quoted by the bidder without price bid in technical cover	
13	All details of the items (HSN, MSME, Make-in-India, make/brand, model, pack size and remark) quoted by the bidder.	
14	Any other document(s) enclosed (To be specified)	

I/We certify that the information furnished above is true and correct. The terms and conditions are acceptable to us and have the authority to bid a tender.

Signature of the owner/
Managing Partner/Director

Date:
Place:

Name:
Seal:

Appendix-A
(To printed on the Firm’s letterhead)
Self-certification format for claiming purchase preference under the “Public Procurement preference to Make in India” order

As per the order issued by

(i) Department of Industrial Policy and Promotion (DIPP) vide No. P-45021/2/2017-BE-II dated 15.06.2017 as further amended by Order No.P-45021/2/2017-B.E.-II dated 28.05.2018, Order No.P-45021/2/2017-B.E.-II dated 29.05.2019, Order No. P-45021/2/2017-PP (BE-II) dated 04.06.2020 and Order No. P-45021/2/2017-PP (BE-II) dated 16.09.2020; and

(ii) Department of Pharmaceuticals vide No. F- 31026/36/2016-MD dated 18.05.2018 and the subsequent orders thereof; The purchaser reserves the right to give preference to the local supplier.

A local supplier (definition of “local supplier” is given in clause 2 of the aforesaid order of DIPP as amended from time to time) has to submit the following along with their e-tender(s) failing which their bid will be evaluated without considering such preference mentioned in the DIPP order dated 15.06.2017 further amended on 28.05.2018, 25.09.2019 and 04.06.2020:

- a. The local supplier at the time of e-tender, bidding or solicitation shall be required to provide self-certification that the item offered meets the minimum local content and shall give details of the location(s) at which the local value addition is made in the format in Annexure A.
- b.

<p>“Certified that the following items quoted has more than 50% or 20% to 50% of value addition in India at the location(s) mentioned against each and is eligible for purchase preference as per the Govt. of India “Public Procurement preference to Make in India” order Dt.15.06.2017 as further amended by Order No.P-45021/2/2017-B.E.-II dated 28.05.2018, Order No.P-45021/2/2017- B.E.-II dated 29.05.2019, and Order No. P-45021/2/2017-PP (BE-II) dated 04.06.2020.</p>				
Item Code	Name of the Item	Percentage of Local content	Location of value addition in India.	Mention whether Class I local supplier or Class II local supplier or non-local Supplier
<p>Authorized Signature: Name: Designation:</p>				

c. In cases of procurement for a value in excess of Rs. 10 Crore. the local supplier shall be required to provide a certificate from the statutory auditor or cost auditor of the company (in the case of companies) or from a practicing cost accountant or practicing chartered accountant (in respect of suppliers other than companies) giving the percentage of local content.

d. Minimum Local Content: ‘Class-I local supplier’ means a supplier or service provider, whose goods, services or works offered for procurement, has local content equal to or more than 50%, as defined

under the Order. “Class-II local supplier” means a supplier or service provider, whose goods, services or works offered for procurement, has local content more than 20% but less than 50%, as defined under this Order.

e. Margin of Purchase Preference: The margin of purchase preference shall be 20%.

f. Manufacture under license/technology collaboration agreements with phased indigenization are exempted from meeting the stipulated local content if the product is being manufactured in India under a license from a foreign manufacturer who holds intellectual property rights and where there is a technology collaboration agreement/transfer of technology agreement for indigenous manufacture of a product developed abroad with clear phasing of increase in local content.

g. Decisions on complaints relating to implementation of this Order shall be taken by the competent authority which is empowered to look into procurement-related complaints relating the procuring entity.

h. A constituted committee with internal and external experts will examine for independent verification of self-declarations and auditor’s/accountant’s certificates on random basis and in the case of complaints.

i. A fees of Rs.10000/- in the form of demand draft favoring The Director, JIPMER, payable at Puducherry, is required to be deposited with complaints for verification of local content.

j. False declarations will be breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which a bidder or its successors can be debarred for up to two years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under law.

k. A supplier who has been debarred by any procuring entity for violation of this Order shall not be eligible for preference under this Order for procurement by any other procuring entity for the duration of the debarment. The debarment for such other procuring entities shall take effect prospectively from the date on which it comes to the notice of other procurement entities.

SECTION – IX

**GFR-144 (xi) compliance certificate
(To be printed on the Firm’s letterhead)**

Tender No:

GFR-144(xi) compliance certificate (as per order F.No. 6/18/2019-PPD, Ministry of Finance, GOI)

Item Code	Name of the Item	I have read the clauses regarding restrictions under GFR144(xi) on procurement from a bidder of a country which shares a land border with India	Signature of Bidder (required in each row failing which that row will not be considered)
		I certify that the Bidder is NOT from such a country for this item.	

I hereby certify that we fulfill all requirement in this regard and is eligible to be considered for the procurement on CPP portal.

Thanking you.

Authorized Signatory